.-online.com antibodies

Datasheet for ABIN456165 MTDH ELISA Kit



Overview

Quantity:	96 tests
Target:	MTDH
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	1.56-100 nM/L
Minimum Detection Limit:	1.56 nM/L
Application:	ELISA

Product Details

Purpose:	This immunoassay kit allows for the use in vitro quantitative determination of human placenta lactogen, HPL concentrations in cell culture supernates, serum, plasma and other biological fluids.
Sample Type:	Cell Culture Supernatant, Plasma, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay recognizes recombinant and natural human HPL.
Cross-Reactivity (Details):	No significant cross-reactivity or interference was observed.
Sensitivity:	< 0.078 ng/mL
	The sensitivity of this assay, or Lower Limit of Detection (LLD) was defined as the lowest
	detectable concentration that could be differentiated from zero.

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/5 | Product datasheet for ABIN456165 | 09/12/2023 | Copyright antibodies-online. All rights reserved.

Product Details	
Characteristics:	Homo sapiens,Human,Protein LYRIC,3D3/LYRIC,Astrocyte elevated gene-1 protein,AEG- 1,Lysine-rich CEACAM1 co-isolated protein,Metadherin,Metastasis adhesion protein,MTDH,AEG1,LYRIC
Components:	Reagent (Quantity): Assay plate (1), Standard (2), Sample Diluent (1x20ml), Assay Diluent A (1x10ml), Assay DiluentB 1 x 10ml Detection Reagent A (1x120µl), Detection Reagent B (1x120µl), Wash Buffer 1 x 30ml 2 (25 x concentrate) Substrate (1x10ml), Stop Solution (1x10ml)

Target Details

Target:	MTDH
Alternative Name:	MTDH (MTDH Products)
Background:	Human placental lactogen (hPL) is a hormone produced by the placenta, the organ that
	develops during pregnancy to help feed the growing baby. This hormone breaks down fats from
	the mother to provide fuel for the the growing baby. It can lead to insulin resistance and
	carbohydrate intolerance in the mother. A test can be done to measure the amount of hPL in
	the blood. The test is only done on pregnant women.
Pathways:	Cellular Response to Molecule of Bacterial Origin, Cell-Cell Junction Organization

Application Details

Sample Volume:	100 µL
Plate:	Pre-coated
Protocol:	The microtiter plate provided in this kit has been pre-coated with an antibody specific to HPL.
	Standards or samples are then added to the appropriate microtiter plate wells with a biotin-
	conjugated polyclonal antibody preparation specific for HPL and Avidin conjugated to
	Horseradish Peroxidase (HRP) is added to each microplate well and incubated. Then a TMB
	(3,3'5, 5' tetramethyl-benzidine) substrate solution is added to each well. Only those wells that
	contain HPL, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change ir
	color. The enzyme-substrate reaction is terminated by the addition of a sulphuric acid solution
	and the color change is measured spectrophotometrically at a wavelength of 450 nm \pm 2 nm.
	The concentration of HPL in the samples is then determined by comparing the O.D. of the
	samples to the standard curve.
Reagent Preparation:	Bring all reagents to room temperature before use. Wash Buffer - If crystals have formed in the

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 2/5 | Product datasheet for ABIN456165 | 09/12/2023 | Copyright antibodies-online. All rights reserved.

Application Details

concentrate, warm to room temperature and mix gently until the crystals have completely dissolved. Dilute 30 mL of Wash Buffer Concentrate into deionized or distilled water to prepare 750 mL of Wash Buffer. Standard - Reconstitute the Standard with 1.0 mL of Sample Diluent. This reconstitution produces a stock solution of 20 ng/mL. Allow the standard to sit for a minimum of 15 minutes with gentle agitation prior to making serial dilutions. The undiluted 3 standard serves as the high standard (20 ng/mL). The Sample Diluent serves as the zero standard (0 ng/mL). Detection Reagent A and B - Dilute to the working concentration specified on the vial label using Assay Diluent A and B (1:100), respectively.

Serum - Use a serum separator tube (SST) and allow samples to clot for 30 minutes before Sample Collection: centrifugation for 15 minutes at approximately 1000 x g. Remove serum and assay immediately or aliquot and store samples at -20 °C or -80 °C. Plasma - Collect plasma using EDTA or heparin as an anticoagulant. Centrifuge samples for 15 minutes at 1000 x g at 2 - 8 °C within 30 minutes of collection. Store samples at -20 °C or -80 °C. Avoid repeated freeze-thaw cycles. Cell culture supernates and other biological fluids - Remove particulates by centrifugation and assay immediately or aliquot and store samples at -20 °C or -80 °C. Avoid repeated freeze-thaw cycles. Sample preparation - Serum/plasma samples require a 1000 fold dilution. A suggested 1000-fold dilution is 10uLsample + 9990uL Sample Diluent. Note: Serum, plasma, and cell culture supernatant samples to be used within 7 days may be stored at 2-8C, otherwise samples must stored at -20 °C (\leq 3 months) or -80 °C (\leq 6 months) to avoid loss of bioactivity and contamination. Avoid freeze-thaw cycles. When performing the assay slowly bring samples to room temperature. It is recommended that all samples be assayed in duplicate. Assay Procedure: Allow all reagents to reach room temperature. All the reagents should be mixed thoroughly by

gently swirling before pipetting. Avoid foaming. Arrange and label required number of strips. Prepare all reagents, working standards and samples as directed in the previous sections. 1. Add 100 uL of Standard, Blank, or Sample per well. Cover with the adhesive strip. Incubate for 2 hours at 37 °C.

2. Remove the liquid of each well, don't wash.

 Add 100 uL of Detection Reagent A working solution to each well. Incubate for 1 hour at 37°C. Detection Reagent A working solution may appear cloudy. Warm to room temperature and mix gently until solution appears uniform.

4. Aspirate each well and wash, repeating the process three times for a total of three washes. Wash by filling each well with Wash Buffer (350 uL) using a squirt bottle, multi-channel pipette, manifold dispenser or autowasher. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 3/5 | Product datasheet for ABIN456165 | 09/12/2023 | Copyright antibodies-online. All rights reserved. decanting. Invert the plate and blot it against clean paper towels.

5. Add 100 uL of Detection Reagent B working solution to each well. Cover with a new adhesive strip.Incubate for 1 hours at 37 °C.

6. Repeat the aspiration/wash as in step

4. 7. Add 90 uL of Substrate Solution to each well. Incubate within 30 minutes at 37°C. Protect	
from light.	

8. Add 50 uL of Stop Solution to each well. If color change does not appear uniform, gently tap the plate to ensure thorough mixing.

9. Determine the optical density of each well at once, using a microplate reader set to 450 nm. Important Note:

 Please carefully reconstitute Standards or working Detection Reagent A and B according to 4 the instruction, and avoid foaming and mix gently until the crystals have completely dissolved. The reconstituted Standards can be used only once.

2. The wash procedure is critical. Insufficient washing will result in poor precision and falsely elevated absorbance readings.

3. It is recommended that no more than 32 wells be used for each assay run if manual pipetting is used since pipetting of all standards, specimens and controls should be completed within 5 minutes. A full plate of 96 wells may be used if automated pipetting is available.

4. Duplication of all standards and specimens, although not required, is recommended.

5. When mixing or reconstituting protein solutions, always avoid foaming.

6. To avoid cross-contamination, change pipette tips between additions of each standard level, between sample additions, and between reagent additions. Also, use separate reservoirs for each reagent.

7. To ensure accurate results, proper adhesion of plate sealers during incubation steps is necessary.

8. Do not substitute reagents from one kit lot to another. Use only the reagents supplied by manufacturer.

Calculation of Results: Average the duplicate readings for each standard, control, and sample and subtract the average zero standard optical density. Create a standard curve by reducing the data using computer software capable of generating a four parameter logistic (4-PL) curve-fit. As an alternative, construct a standard curve by plotting the mean absorbance for each standard on the y-axis against the concentration on the x-axis and draw a best fit curve through the points on the graph. The data may be linearized by plotting the log of the HPL concentrations versus the log of the O.D. and the best fit line can be determined by regression analysis. This procedure will produce an adequate but less precise fit of the data. If samples have been diluted, the

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 4/5 | Product datasheet for ABIN456165 | 09/12/2023 | Copyright antibodies-online. All rights reserved.

Application Details	
	concentration read from the standard curve must be multiplied by the dilution factor.
Restrictions:	For Research Use only
Handling	
Handling Advice:	1. The kit should not be used beyond the expiration date on the kit label.
	2. Do not mix or substitute reagents with those from other lots or sources.
	3. If samples generate values higher than the highest standard, further dilute the samples with
	the Assay Diluent and repeat the assay. Any variation in standard diluent, operator, pipetting
	technique, washing technique,incubation time or temperature, and kit age can cause variation ir
	binding.
	4. This assay is designed to eliminate interference by soluble receptors, ligands, binding
	proteins, and other factors present in biological samples. Until all factors have been tested in
	the Immunoassay, the possibility of interference cannot be excluded.
Storage:	4 °C/-20 °C
Storage Comment:	The Standard, Detection Reagent A, Detection Reagent B and the 96-well strip plate should be stored at -20 °C upon being received. The other reagents can be stored at 4 °C.