



[Go to Product page](#)

Datasheet for ABIN4881333
Substance P ELISA Kit

Overview

Quantity:	96 tests
Target:	Substance P
Reactivity:	Human, Mouse, Rat
Method Type:	Competition ELISA
Detection Range:	0.1-1.000 ng/mL
Minimum Detection Limit:	0.1 ng/mL
Application:	ELISA

Product Details

Purpose:	Human/Mouse/Rat Substance P EIA Kit optimized for serum, plasma, and cell culture supernatants. Competition-based ELISA on a 96-well strip plate.
Sample Type:	Plasma, Cell Culture Supernatant, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This EIA kit is designed to detect human, mouse, and rat Substance P.
Characteristics:	<ul style="list-style-type: none">• Strip plates and additional reagents allow for use in multiple experiments• Quantitative protein detection• Establishes normal range• The best products for confirmation of antibody array data
Components:	<ul style="list-style-type: none">• Pre-Coated 96-well Strip Microplate• Wash Buffer

Product Details

- Standard Peptide
- Assay Diluent(s)
- Biotinylated Peptide
- HRP-Streptavidin
- TMB One-Step Substrate
- Stop Solution
- Assay Diagram
- Positive Control Sample
- Capture Antibody
- User Manual

Material not included:

- Distilled or deionized water
- Precision pipettes to deliver 2 μ L to 1 mL volumes
- Adjustable 1-25 mL pipettes for reagent preparation
- 100 mL and 1 liter graduated cylinders
- Tubes to prepare standard and sample dilutions
- Orbital shaker
- Aluminum foil
- Saran Wrap
- Absorbent paper
- Microplate reader capable of measuring absorbance at 450nm
- SigmaPlot software (or other software that can perform four-parameter logistic regression models)

Target Details

Target: Substance P

Abstract: [Substance P Products](#)

Gene ID: 6863

UniProt: [P20366](#)

Application Details

Application Notes: Recommended Dilution for serum and plasma samples Human: 2x / Mouse: 8x / Rat: 8X

Sample Volume: 100 μ L

Plate: Pre-coated

Protocol:

1. Prepare all reagents, samples and standards as instructed.
2. Add 100 μ L detection antibody to each well.
3. Incubate 1.5 h at RT or O/N at 4 $^{\circ}$ C.

Application Details

4. Add 100 μ L standard or sample to each well.
5. Incubate 2.5 h at RT.
6. Add 100 μ L prepared streptavidin solution.
7. Incubate 45 min at RT.
8. Add 100 μ L TMB One-Step Substrate Reagent to each well.
9. Incubate 30 min at RT.
10. Add 50 μ L Stop Solution to each well.
11. Read plate at 450 nm immediately.

Restrictions: For Research Use only

Handling

Storage: -20 °C

Storage Comment: Standard, biotinylated peptide, and positive control should be stored at -20°C after arrival. Avoid multiple freeze-thaws. The remaining kit components may be stored at 4°C. Opened microplate wells and antibody (Item N) may be stored for up to 1 month at 2° to 8°C. Return unused wells to the pouch containing desiccant pack and reseal along entire edge.

Expiry Date: 6 months