

[Go to Product page](#)

## Datasheet for ABIN4881336 Urocortin ELISA Kit

### Overview

Quantity:	96 tests
Target:	Urocortin (UCN)
Reactivity:	Mouse, Human, Rat
Method Type:	Competition ELISA
Detection Range:	0.1-1.000 ng/mL
Minimum Detection Limit:	0.1 ng/mL
Application:	ELISA

### Product Details

Purpose:	Human/Mouse/Rat Urocortin EIA Kit optimized for serum. Competition-based ELISA on a 96-well strip plate.
Sample Type:	Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This EIA kit is designed to detect human, mouse, and rat active Urocortin.
Sensitivity:	1.2 ng/mL
Characteristics:	<ul style="list-style-type: none"><li>• Strip plates and additional reagents allow for use in multiple experiments</li><li>• Quantitative protein detection</li><li>• Establishes normal range</li><li>• The best products for confirmation of antibody array data</li></ul>

## Product Details

Components:	<ul style="list-style-type: none"><li>• Pre-Coated 96-well Strip Microplate</li><li>• Wash Buffer</li><li>• Standard Peptide</li><li>• Assay Diluent(s)</li><li>• Biotinylated Peptide</li><li>• HRP-Streptavidin</li><li>• TMB One-Step Substrate</li><li>• Stop Solution</li><li>• Assay Diagram</li><li>• Positive Control Sample</li><li>• Capture Antibody</li><li>• User Manual</li></ul>
-------------	---

Material not included:	<ul style="list-style-type: none"><li>• Distilled or deionized water</li><li>• Precision pipettes to deliver 2 µL to 1 mL volumes</li><li>• Adjustable 1-25 mL pipettes for reagent preparation</li><li>• 100 mL and 1 liter graduated cylinders</li><li>• Tubes to prepare standard and sample dilutions</li><li>• Orbital shaker</li><li>• Aluminum foil</li><li>• Saran Wrap</li><li>• Absorbent paper</li><li>• Microplate reader capable of measuring absorbance at 450nm</li><li>• SigmaPlot software (or other software that can perform four-parameter logistic regression models)</li></ul>
------------------------	--

## Target Details

Target:	Urocortin (UCN)
Alternative Name:	Urocortin ( <a href="#">UCN Products</a> )
Gene ID:	7349
UniProt:	<a href="#">P55089</a>
Pathways:	<a href="#">Positive Regulation of Peptide Hormone Secretion</a> , <a href="#">Hormone Activity</a> , <a href="#">Negative Regulation of Hormone Secretion</a> , <a href="#">cAMP Metabolic Process</a> , <a href="#">Regulation of Cell Size</a> , <a href="#">Feeding Behaviour</a>

## Application Details

Sample Volume:	100 µL
Plate:	Pre-coated

## Application Details

---

Protocol:	<ol style="list-style-type: none"><li>1. Prepare all reagents, samples and standards as instructed.</li><li>2. Add 100 µL detection antibody to each well.</li><li>3. Incubate 1.5 h at RT or O/N at 4 °C.</li><li>4. Add 100 µL standard or sample to each well.</li><li>5. Incubate 2.5 h at RT.</li><li>6. Add 100 µL prepared streptavidin solution.</li><li>7. Incubate 45 min at RT.</li><li>8. Add 100 µL TMB One-Step Substrate Reagent to each well.</li><li>9. Incubate 30 min at RT.</li><li>10. Add 50 µL Stop Solution to each well.</li><li>11. Read plate at 450 nm immediately.</li></ol>
-----------	---

---

Restrictions:	For Research Use only
---------------	-----------------------

---

## Handling

---

Storage:	-20 °C
Storage Comment:	<p>Standard, Biotinylated Urocortin peptide, and Positive Control should be stored at -20°C after arrival. Avoid multiple freeze-thaws. The remaining kit components may be stored at 4°C.</p> <p>Opened Microplate Wells and antibody (Item N) may be stored for up to 1 month at 2° to 8°C.</p> <p>Return unused wells to the pouch containing desiccant pack and reseal along entire edge.</p>
Expiry Date:	6 months

---