

Datasheet for ABIN4881517

c-MET ELISA Kit

[Go to Product page](#)

Overview

Quantity:	96 tests
Target:	c-MET (MET)
Reactivity:	Dog
Method Type:	Sandwich ELISA
Application:	ELISA

Product Details

Purpose:	Custom Canine (dog) Hepatocyte growth factor receptor ELISA Kit.
Sample Type:	Cell Culture Supernatant, Cell Lysate, Plasma, Serum, Tissue Lysate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	The antibody pair provided in this kit recognizes Canine Hepatocyte growth factor receptor.
Characteristics:	• Strip plates and additional reagents allow for use in multiple experiments • Quantitative protein detection • Establishes normal range • The best products for confirmation of antibody array data
Components:	• Pre-Coated 96-well Strip Microplate • Wash Buffer • Stop Solution • Assay Diluent(s) • Lyophilized Standard • Biotinylated Detection Antibody

Product Details

- Streptavidin-Conjugated HRP
- TMB One-Step Substrate

Material not included:

- Distilled or deionized water
- Precision pipettes to deliver 2 µL to 1 µL volumes
- Adjustable 1-25 µL pipettes for reagent preparation
- 100 µL and 1 liter graduated cylinders
- Tubes to prepare standard and sample dilutions
- Absorbent paper
- Microplate reader capable of measuring absorbance at 450nm
- Log-log graph paper or computer and software for ELISA data analysis

Target Details

Target:	c-MET (MET)
Alternative Name:	HGFR (MET Products)
Gene ID:	403438
UniProt:	Q75ZY9
Pathways:	RTK Signaling , Carbohydrate Homeostasis , Synaptic Membrane , Signaling of Hepatocyte Growth Factor Receptor

Application Details

Sample Volume:	100 µL
Plate:	Pre-coated
Protocol:	1. Prepare all reagents, samples and standards as instructed in the manual. 2. Add 100 µL of standard or sample to each well. 3. Incubate 2.5 h at RT or O/N at 4 °C. 4. Add 100 µL of prepared biotin antibody to each well. 5. Incubate 1 h at RT. 6. Add 100 µL of prepared Streptavidin solution to each well. 7. Incubate 45 min at RT. 8. Add 100 µL of TMB One-Step Substrate Reagent to each well. 9. Incubate 30 min at RT. 10. Add 50 µL of Stop Solution to each well. 11. Read at 450 nm immediately.
Restrictions:	For Research Use only

Handling

Storage:	-20 °C
Storage Comment:	The entire kit may be stored at -20°C for up to 1 year from the date of shipment. Avoid repeated freeze-thaw cycles. The kit may be stored at 4°C for up to 6 months. For extended storage, it is recommended to store at -80°C.
Expiry Date:	6 months