



[Go to Product page](#)

Datasheet for ABIN4881926  
**CD276 ELISA Kit**

### Overview

Quantity: 96 tests

Target: CD276

Reactivity: Human

Method Type: Sandwich ELISA

Application: ELISA

### Product Details

Purpose: Custom Human B7 Homolog 3 (CD276) ELISA Kit.

Sample Type: Cell Culture Supernatant, Cell Lysate, Plasma, Serum, Tissue Lysate

Analytical Method: Quantitative

Detection Method: Colorimetric

Specificity: The antibody pair provided in this kit recognizes Human B7 Homolog 3 (CD276)

Characteristics:

- Strip plates and additional reagents allow for use in multiple experiments
- Quantitative protein detection
- Establishes normal range
- The best products for confirmation of antibody array data

Components:

- Pre-Coated 96-well Strip Microplate
- Wash Buffer
- Stop Solution
- Assay Diluent(s)
- Lyophilized Standard
- Biotinylated Detection Antibody

## Product Details

---

- Streptavidin-Conjugated HRP
- TMB One-Step Substrate

---

### Material not included:

- Distilled or deionized water
- Precision pipettes to deliver 2 µL to 1 µL volumes
- Adjustable 1-25 µL pipettes for reagent preparation
- 100 µL and 1 liter graduated cylinders
- Tubes to prepare standard and sample dilutions
- Absorbent paper
- Microplate reader capable of measuring absorbance at 450nm
- Log-log graph paper or computer and software for ELISA data analysis

## Target Details

---

Target: CD276

Alternative Name: B7-H3 ([CD276 Products](#))

Gene ID: 80381

UniProt: [Q5ZPR3](#)

Pathways: [Cancer Immune Checkpoints](#)

## Application Details

---

Sample Volume: 100 µL

Plate: Pre-coated

Protocol:

1. Prepare all reagents, samples and standards as instructed in the manual.
2. Add 100 µL of standard or sample to each well.
3. Incubate 2.5 h at RT or O/N at 4 °C.
4. Add 100 µL of prepared biotin antibody to each well.
5. Incubate 1 h at RT.
6. Add 100 µL of prepared Streptavidin solution to each well.
7. Incubate 45 min at RT.
8. Add 100 µL of TMB One-Step Substrate Reagent to each well.
9. Incubate 30 min at RT.
10. Add 50 µL of Stop Solution to each well.
11. Read at 450 nm immediately.

---

Restrictions: For Research Use only

## Handling

---

Storage: -20 °C

---

Storage Comment: The entire kit may be stored at -20°C for up to 1 year from the date of shipment. Avoid repeated freeze-thaw cycles. The kit may be stored at 4°C for up to 6 months. For extended storage, it is recommended to store at -80°C.

---

Expiry Date: 6 months