.-online.com antibodies

# Datasheet for ABIN4882138 Poliovirus Receptor ELISA Kit



#### Overview

| Quantity:    | 96 tests                  |
|--------------|---------------------------|
| Target:      | Poliovirus Receptor (PVR) |
| Reactivity:  | Human                     |
| Method Type: | Sandwich ELISA            |
| Application: | ELISA                     |

#### **Product Details**

| Purpose:           | Human CD155/PVR ELISA Kit for Serum, Plasma, and Cell Culture Supernatants.   |
|--------------------|---|
| Sample Type:       | Plasma, Cell Culture Supernatant, Serum   |
| Analytical Method: | Quantitative  |
| Detection Method:  | Colorimetric  |
| Specificity:       | This ELISA antibody pair recognizes Human CD155.  |
| Characteristics:   | <ul> <li>Strip plates and additional reagents allow for use in multiple experiments</li> <li>Quantitative protein detection</li> <li>Establishes normal range</li> <li>The best products for confirmation of antibody array data</li> </ul> |
| Components:        | <ul> <li>Pre-Coated 96-well Strip Microplate</li> <li>Wash Buffer</li> <li>Stop Solution</li> <li>Assay Diluent(s)</li> <li>Lyophilized Standard</li> <li>Biotinylated Detection Antibody</li> </ul>  |

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/3 | Product datasheet for ABIN4882138 | 09/11/2023 | Copyright antibodies-online. All rights reserved.

|                        | <ul><li>Streptavidin-Conjugated HRP</li><li>TMB One-Step Substrate</li></ul> |
|------------------------|--|
| Material not included: | Distilled or deionized water   |
|                        | - Precision pipettes to deliver 2 $\mu$ L to 1 $\mu$ L volumes               |
|                        | <ul> <li>Adjustable 1-25 µL pipettes for reagent preparation</li> </ul>      |
|                        | <ul> <li>100 µL and 1 liter graduated cylinders</li> </ul>                   |
|                        | Tubes to prepare standard and sample dilutions                               |
|                        | Absorbent paper  |
|                        | Microplate reader capable of measuring absorbance at 450nm                   |
|                        | Log-log graph paper or computer and software for ELISA data analysis         |

### Target Details

| Target:           | Poliovirus Receptor (PVR)  |
|-------------------|--|
| Alternative Name: | CD155 (PVR Products)   |
| Gene ID:          | 5817   |
| UniProt:          | P15151   |
| Pathways:         | Regulation of Leukocyte Mediated Immunity, Positive Regulation of Immune Effector Process, |
|                   | Cell-Cell Junction Organization, Cancer Immune Checkpoints, SARS-CoV-2 Protein Interactome |

## Application Details

| Sample Volume: | 100 µL  |
|----------------|---|
| Plate:         | Pre-coated  |
| Protocol:      | 1. Prepare all reagents, samples and standards as instructed in the manual. |
|                | 2. Add 100 µL of standard or sample to each well.                           |
|                | 3. Incubate 2.5 h at RT or O/N at 4 °C.                                     |
|                | 4. Add 100 $\mu$ L of prepared biotin antibody to each well.                |
|                | 5. Incubate 1 h at RT.  |
|                | 6. Add 100 $\mu$ L of prepared Streptavidin solution to each well.          |
|                | 7. Incubate 45 min at RT.   |
|                | 8. Add 100 µL of TMB One-Step Substrate Reagent to each well.               |
|                | 9. Incubate 30 min at RT.   |
|                | 10. Add 50 µL of Stop Solution to each well.                                |
|                | 11. Read at 450 nm immediately.   |

Restrictions:

For Research Use only

### Handling

| Storage:         | -20 °C   |
|------------------|--|
| Storage Comment: | The entire kit may be stored at -20°C for up to 1 year from the date of shipment. Avoid repeated freeze-thaw cycles. The kit may be stored at 4°C for up to 6 months. For extended storage, it is recommended to store at -80°C. |
| Expiry Date:     | 6 months   |