

Datasheet for ABIN4882265

CD86 ELISA Kit

1 Image



Overview

Quantity:	96 tests
Target:	CD86
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.82-200 ng/mL
Minimum Detection Limit:	0.82 ng/mL
Application:	ELISA

Product Details

Product Details	
Purpose:	Human CD86 (B7-2) ELISA Kit for Cell Culture Supernatants, Plasma, and Serum samples
Sample Type:	Plasma, Cell Culture Supernatant, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	The antibody pair provided in this kit recognizes Human CD86 (B7-2)
Characteristics:	Strip plates and additional reagents allow for use in multiple experiments
	Quantitative protein detection
	Establishes normal range
	The best products for confirmation of antibody array data
Components:	Pre-Coated 96-well Strip Microplate
	Wash Buffer
	Stop Solution

Product Details

- Assay Diluent(s)
- · Lyophilized Standard
- · Biotinylated Detection Antibody
- · Streptavidin-Conjugated HRP
- · TMB One-Step Substrate

Material not included:

- Distilled or deionized water
- Precision pipettes to deliver 2 μL to 1 μL volumes
- Adjustable 1-25 µL pipettes for reagent preparation
- 100 µL and 1 liter graduated cylinders
- · Tubes to prepare standard and sample dilutions
- · Absorbent paper
- Microplate reader capable of measuring absorbance at 450nm
- · Log-log graph paper or computer and software for ELISA data analysis

Target Details

Target:	CD86
Alternative Name:	CD86 (CD86 Products)
Gene ID:	942
UniProt:	P42081
Pathways:	TCR Signaling, Fc-epsilon Receptor Signaling Pathway, EGFR Signaling Pathway, Neurotrophin
	Signaling Pathway, Activation of Innate immune Response, Cellular Response to Molecule of
	Bacterial Origin, Positive Regulation of Immune Effector Process, Activated T Cell Proliferation

Application Details

Application Notes:	Recommended Dilution for serum and plasma samples2 fold
Sample Volume:	100 μL
Plate:	Pre-coated Pre-coated
Protocol:	 Prepare all reagents, samples and standards as instructed in the manual. Add 100 μL of standard or sample to each well. Incubate 2.5 h at RT or O/N at 4 °C. Add 100 μL of prepared biotin antibody to each well. Incubate 1 h at RT. Add 100 μL of prepared Streptavidin solution to each well. Incubate 45 min at RT.

8. Add 100 μL of TMB One-Step Substrate Reagent to each well.

9. Incubate 30 min at RT.

10. Add 50 µL of Stop Solution to each well.

11. Read at 450 nm immediately.

Restrictions:

For Research Use only

6 months


Handling

Storage: -20 °C

Storage Comment: The entire kit may be stored at -20 °C for up to 1 year from the date of shipment. Avoid repeated freeze-thaw cycles. The kit may be stored at 4 °C for up to 6 months. For extended storage, it is recommended to store at -80 °C.

Images

Expiry Date:



ELISA

Image 1.

Human CD86/B7-2 concentration (ng/ml)