

## Datasheet for ABIN4882330 Chitotriosidase 1 ELISA Kit



Overview

Quantity:	96 tests
Target:	Chitotriosidase 1 (CHIT1)
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	8-2.000 pg/mL
Minimum Detection Limit:	8 pg/mL
Application:	ELISA

## Product Details

Purpose:	Human Chitotriosidase ELISA Kit for Serum, Plasma, and Cell Culture Supernatants.
Sample Type:	Plasma, Cell Culture Supernatant, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This ELISA antibody pair recognizes Human Chitotriosidase.
Characteristics:	<ul> <li>Strip plates and additional reagents allow for use in multiple experiments</li> <li>Quantitative protein detection</li> <li>Establishes normal range</li> <li>The best products for confirmation of antibody array data</li> </ul>
Components:	<ul> <li>Pre-Coated 96-well Strip Microplate</li> <li>Wash Buffer</li> <li>Stop Solution</li> </ul>

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/3 | Product datasheet for ABIN4882330 | 07/25/2024 | Copyright antibodies-online. All rights reserved.

	<ul> <li>Assay Diluent(s)</li> <li>Lyophilized Standard</li> <li>Biotinylated Detection Antibody</li> <li>Streptavidin-Conjugated HRP</li> <li>TMB One-Step Substrate</li> </ul>
Material not included:	<ul> <li>Distilled or deionized water</li> <li>Precision pipettes to deliver 2 µL to 1 µL volumes</li> <li>Adjustable 1-25 µL pipettes for reagent preparation</li> <li>100 µL and 1 liter graduated cylinders</li> <li>Tubes to prepare standard and sample dilutions</li> <li>Absorbent paper</li> <li>Microplate reader capable of measuring absorbance at 450nm</li> <li>Log-log graph paper or computer and software for ELISA data analysis</li> </ul>

## Target Details

Target:	Chitotriosidase 1 (CHIT1)
Alternative Name:	Chitotriosidase (CHIT1 Products)
Gene ID:	1118
UniProt:	Q13231

## Application Details

Application Notes:	Recommended Dilution for serum and plasma samples80 fold
Sample Volume:	100 µL
Plate:	Pre-coated
Protocol:	1. Prepare all reagents, samples and standards as instructed in the manual.
	2. Add 100 $\mu$ L of standard or sample to each well.
	3. Incubate 2.5 h at RT or O/N at 4 °C.
	4. Add 100 $\mu$ L of prepared biotin antibody to each well.
	5. Incubate 1 h at RT.
	6. Add 100 $\mu$ L of prepared Streptavidin solution to each well.
	7. Incubate 45 min at RT.
	8. Add 100 $\mu$ L of TMB One-Step Substrate Reagent to each well.
	9. Incubate 30 min at RT.
	10. Add 50 μL of Stop Solution to each well.
	11. Read at 450 nm immediately.

Application Details	
Restrictions:	For Research Use only
Handling	
Storage:	-20 °C
Storage Comment:	The entire kit may be stored at -20°C for up to 1 year from the date of shipment. Avoid repeated freeze-thaw cycles. The kit may be stored at 4°C for up to 6 months. For extended storage, it is recommended to store at -80°C.
Expiry Date:	6 months