.-online.com antibodies

Datasheet for ABIN4882357 CILP ELISA Kit



Overview

Quantity:	96 tests
Target:	CILP
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.053-13 ng/mL
Minimum Detection Limit:	0.053 ng/mL
Application:	ELISA

Product Details

Purpose:	Human CILP-1 ELISA Kit for Serum, Plasma, and Cell Culture Supernatants.
Sample Type:	Plasma, Cell Culture Supernatant, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This ELISA antibody pair recognizes Human CILP-1.
Sensitivity:	0.05 ng/mL
Characteristics:	 Strip plates and additional reagents allow for use in multiple experiments Quantitative protein detection Establishes normal range The best products for confirmation of antibody array data
Components:	Pre-Coated 96-well Strip Microplate

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/3 | Product datasheet for ABIN4882357 | 09/11/2023 | Copyright antibodies-online. All rights reserved.

	Wash Buffer
	Stop Solution
	Assay Diluent(s)
	Lyophilized Standard
	Biotinylated Detection Antibody
	Streptavidin-Conjugated HRP
	TMB One-Step Substrate
Material not included:	Distilled or deionized water
	• Precision pipettes to deliver 2 μ L to 1 μ L volumes
	 Adjustable 1-25 µL pipettes for reagent preparation
	 100 μL and 1 liter graduated cylinders
	Tubes to prepare standard and sample dilutions
	Absorbent paper
	Microplate reader capable of measuring absorbance at 450nm

• Log-log graph paper or computer and software for ELISA data analysis

Target Details

Target:	CILP
Alternative Name:	CILP-1 (CILP Products)
Gene ID:	8483
UniProt:	075339

Application Details

Application Notes:	Recommended Dilution for serum and plasma samples2 fold
Sample Volume:	100 μL
Plate:	Pre-coated
Protocol:	1. Prepare all reagents, samples and standards as instructed in the manual.
	2. Add 100 μ L of standard or sample to each well.
	3. Incubate 2.5 h at RT or O/N at 4 °C.
	4. Add 100 μ L of prepared biotin antibody to each well.
	5. Incubate 1 h at RT.
	6. Add 100 μ L of prepared Streptavidin solution to each well.
	7. Incubate 45 min at RT.
	8. Add 100 μ L of TMB One-Step Substrate Reagent to each well.
	9. Incubate 30 min at RT.
	10. Add 50 μ L of Stop Solution to each well.

Application Details	
	11. Read at 450 nm immediately.
Restrictions:	For Research Use only
Handling	
Storage:	-20 °C
Storage: Storage Comment:	-20 °C The entire kit may be stored at -20°C for up to 1 year from the date of shipment. Avoid repeated freeze-thaw cycles. The kit may be stored at 4°C for up to 6 months. For extended storage, it is recommended to store at -80°C.