Datasheet for ABIN4882366

Galectin 10 ELISA Kit

Go to Product page

Overview

Quantity:	96 tests
Target:	Galectin 10 (CLC)
Reactivity:	Human
Method Type:	Sandwich ELISA
Application:	ELISA

Product Details

Purpose:	Custom Human CLC ELISA Kit.
Sample Type:	Cell Culture Supernatant, Cell Lysate, Plasma, Serum, Tissue Lysate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This ELISA antibody pair recognizes Human CLC. Other species not yet determined.
Characteristics:	• Strip plates and additional reagents allow for use in multiple experiments • Quantitative protein detection • Establishes normal range • The best products for confirmation of antibody array data
Components:	• Pre-Coated 96-well Strip Microplate • Wash Buffer • Stop Solution • Assay Diluent(s) • Lyophilized Standard • Biotinylated Detection Antibody

Product Details

	• Streptavidin-Conjugated HRP • TMB One-Step Substrate
Material not included:	• Distilled or deionized water • Precision pipettes to deliver 2 µL to 1 µL volumes • Adjustable 1-25 µL pipettes for reagent preparation • 100 µL and 1 liter graduated cylinders • Tubes to prepare standard and sample dilutions • Absorbent paper • Microplate reader capable of measuring absorbance at 450nm • Log-log graph paper or computer and software for ELISA data analysis

Target Details

Target:	Galectin 10 (CLC)
Alternative Name:	CLC (CLC Products)
Gene ID:	1178
UniProt:	Q05315
Pathways:	Regulation of Leukocyte Mediated Immunity, Positive Regulation of Immune Effector Process, Production of Molecular Mediator of Immune Response

Application Details

Sample Volume:	100 µL
Plate:	Pre-coated
Protocol:	1. Prepare all reagents, samples and standards as instructed in the manual. 2. Add 100 µL of standard or sample to each well. 3. Incubate 2.5 h at RT or O/N at 4 °C. 4. Add 100 µL of prepared biotin antibody to each well. 5. Incubate 1 h at RT. 6. Add 100 µL of prepared Streptavidin solution to each well. 7. Incubate 45 min at RT. 8. Add 100 µL of TMB One-Step Substrate Reagent to each well. 9. Incubate 30 min at RT. 10. Add 50 µL of Stop Solution to each well. 11. Read at 450 nm immediately.
Restrictions:	For Research Use only

Handling

Storage:	-20 °C
Storage Comment:	The entire kit may be stored at -20°C for up to 1 year from the date of shipment. Avoid repeated freeze-thaw cycles. The kit may be stored at 4°C for up to 6 months. For extended storage, it is recommended to store at -80°C.
Expiry Date:	6 months