# antibodies -online.com





# Datasheet for ABIN4882399

# c-MYC ELISA Kit



#### Overview

Quantity:	96 tests
Target:	c-MYC (MYC)
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.62-150 ng/mL
Minimum Detection Limit:	0.62 ng/mL
Application:	ELISA

#### Product Datails

Product Details	
Purpose:	Human c-Myc (MYC) ELISA Kit for serum, plasma, and cell culture supernatant samples.
Sample Type:	Plasma, Cell Culture Supernatant, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This ELISA antibody pair detects human c-Myc. Other species not determined.
Characteristics:	Strip plates and additional reagents allow for use in multiple experiments
	Quantitative protein detection
	Establishes normal range
	The best products for confirmation of antibody array data
Components:	Pre-Coated 96-well Strip Microplate
	Wash Buffer
	Stop Solution

#### **Product Details**

- Assay Diluent(s)
- · Lyophilized Standard
- · Biotinylated Detection Antibody
- · Streptavidin-Conjugated HRP
- · TMB One-Step Substrate

#### Material not included:

- Distilled or deionized water
- Precision pipettes to deliver 2  $\mu$ L to 1  $\mu$ L volumes
- Adjustable 1-25 µL pipettes for reagent preparation
- 100 µL and 1 liter graduated cylinders
- · Tubes to prepare standard and sample dilutions
- · Absorbent paper
- Microplate reader capable of measuring absorbance at 450nm
- · Log-log graph paper or computer and software for ELISA data analysis

## **Target Details**

Target:	c-MYC (MYC)
Alternative Name:	C-Myc///MYC (MYC Products)
Gene ID:	4609
UniProt:	P01106
Pathways:	p53 Signaling, Cell Division Cycle, Sensory Perception of Sound, Transition Metal Ion Homeostasis, Mitotic G1-G1/S Phases, Positive Regulation of Endopeptidase Activity, Regulation of Carbohydrate Metabolic Process, Positive Regulation of Response to DNA Damage Stimulus, Warburg Effect

## **Application Details**

Recommended Dilution for serum and plasma samples2 fold
100 μL
Pre-coated
1. Prepare all reagents, samples and standards as instructed in the manual.
2. Add 100 μL of standard or sample to each well.
3. Incubate 2.5 h at RT or O/N at 4 °C.
4. Add 100 $\mu$ L of prepared biotin antibody to each well.
5. Incubate 1 h at RT.
6. Add 100 μL of prepared Streptavidin solution to each well.

# **Application Details**

	7. Incubate 45 min at RT. 8. Add 100 µL of TMB One-Step Substrate Reagent to each well.
	9. Incubate 30 min at RT.
	10. Add 50 μL of Stop Solution to each well.
	11. Read at 450 nm immediately.
Restrictions:	For Research Use only
Handling	
Storage:	-20 °C
Storage Comment:	The entire kit may be stored at -20°C for up to 1 year from the date of shipment. Avoid repeated
	freeze-thaw cycles. The kit may be stored at 4°C for up to 6 months. For extended storage, it is
	recommended to store at -80°C.
Expiry Date:	6 months