

Datasheet for ABIN4883472

ITGA2 ELISA Kit



Overview

Quantity:	96 tests
Target:	ITGA2
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.102 g/mL-25 ng/mL
Minimum Detection Limit:	0.102 ng/mL
Application:	ELISA
Product Details	
Purpose:	Human Integrin alpha 2 (CD49b) ELISA Kit for serum, plasma, and cell culture supernatant samples.
Sample Type:	Plasma, Cell Culture Supernatant, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This ELISA antibody pair detects human ITGA2. Other species not determined.
Sensitivity:	0.10 ng/mL
Characteristics:	 Strip plates and additional reagents allow for use in multiple experiments Quantitative protein detection Establishes normal range The best products for confirmation of antibody array data

Product Details

Components:

- · Pre-Coated 96-well Strip Microplate
- · Wash Buffer
- · Stop Solution
- Assay Diluent(s)
- · Lyophilized Standard
- · Biotinylated Detection Antibody
- · Streptavidin-Conjugated HRP
- TMB One-Step Substrate

Material not included:

- · Distilled or deionized water
- Precision pipettes to deliver 2 μL to 1 μL volumes
- Adjustable 1-25 µL pipettes for reagent preparation
- 100 µL and 1 liter graduated cylinders
- Tubes to prepare standard and sample dilutions
- · Absorbent paper
- Microplate reader capable of measuring absorbance at 450nm
- · Log-log graph paper or computer and software for ELISA data analysis

Target Details

Target:	ITGA2
Alternative Name:	Integrin alpha 2///CD49b (ITGA2 Products)
Gene ID:	3673
UniProt:	P17301
Pathways:	CXCR4-mediated Signaling Events, Smooth Muscle Cell Migration, Integrin Complex

Application Details	
Application Notes:	Recommended Dilution for serum and plasma samples30 fold
Sample Volume:	100 μL
Plate:	Pre-coated
Protocol:	1. Prepare all reagents, samples and standards as instructed in the manual.
	2. Add 100 μL of standard or sample to each well.
	3. Incubate 2.5 h at RT or O/N at 4 °C.
	4. Add 100 μ L of prepared biotin antibody to each well.
	5. Incubate 1 h at RT.
	6. Add 100 μL of prepared Streptavidin solution to each well.
	7. Incubate 45 min at RT.

Application Details

	 8. Add 100 μL of TMB One-Step Substrate Reagent to each well. 9. Incubate 30 min at RT. 10. Add 50 μL of Stop Solution to each well. 11. Read at 450 nm immediately.
Restrictions:	For Research Use only
Handling	
Storage:	-20 °C
Storage Comment:	The entire kit may be stored at -20°C for up to 1 year from the date of shipment. Avoid repeated freeze-thaw cycles. The kit may be stored at 4°C for up to 6 months. For extended storage, it is recommended to store at -80°C.
Expiry Date:	6 months