

Datasheet for ABIN4885058

WNT4 ELISA Kit





Overview

Quantity:	96 tests
Target:	WNT4
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	1.22-300 ng/mL
Minimum Detection Limit:	1.22 ng/mL
Application:	ELISA

Product Details

Product Details	
Purpose:	Human Wnt-4 ELISA Kit for serum, plasma, and cell culture supernatant samples.
Sample Type:	Plasma, Cell Culture Supernatant, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This ELISA antibody pair detects human Wnt-4. Other species not determined.
Characteristics:	 Strip plates and additional reagents allow for use in multiple experiments Quantitative protein detection Establishes normal range The best products for confirmation of antibody array data
Components:	Pre-Coated 96-well Strip MicroplateWash BufferStop Solution

Product Details

- Assay Diluent(s)
- · Lyophilized Standard
- · Biotinylated Detection Antibody
- · Streptavidin-Conjugated HRP
- · TMB One-Step Substrate

Material not included:

- · Distilled or deionized water
- Precision pipettes to deliver 2 μL to 1 μL volumes
- Adjustable 1-25 µL pipettes for reagent preparation
- 100 µL and 1 liter graduated cylinders
- · Tubes to prepare standard and sample dilutions
- · Absorbent paper
- Microplate reader capable of measuring absorbance at 450nm
- · Log-log graph paper or computer and software for ELISA data analysis

Target Details

Target:	WNT4
Alternative Name:	Wnt-4 (WNT4 Products)
Gene ID:	54361
UniProt:	P56705
Pathways:	WNT Signaling, Regulation of Hormone Metabolic Process, Regulation of Hormone Biosynthetic Process, Cell-Cell Junction Organization, Tube Formation

Application Details

Application Notes:	Recommended Dilution for serum and plasma samples2 fold
Sample Volume:	100 μL
Plate:	Pre-coated
Protocol:	1. Prepare all reagents, samples and standards as instructed in the manual.
	2. Add 100 μL of standard or sample to each well.
	3. Incubate 2.5 h at RT or O/N at 4 °C.
	4. Add 100 μL of prepared biotin antibody to each well.
	5. Incubate 1 h at RT.
	6. Add 100 μL of prepared Streptavidin solution to each well.
	7. Incubate 45 min at RT.
	8. Add 100 μL of TMB One-Step Substrate Reagent to each well.
	9. Incubate 30 min at RT.

10. Add 50 μL of Stop Solution to each well.

11. Read at 450 nm immediately.


Restrictions:

For Research Use only

Handling

Storage:	-20 °C
Storage Comment:	The entire kit may be stored at -20°C for up to 1 year from the date of shipment. Avoid repeated freeze-thaw cycles. The kit may be stored at 4°C for up to 6 months. For extended storage, it is recommended to store at -80°C.
Expiry Date:	6 months

Images



ELISA

Image 1.