



[Go to Product page](#)

Datasheet for ABIN4947580
FGF22 ELISA Kit

Overview

Quantity: 96 tests

Target: FGF22

Reactivity: Mouse

Method Type: Sandwich ELISA

Application: ELISA

Product Details

Purpose: Mini Samples ELISA Kit for Fibroblast Growth Factor 22 (FGF22)

Analytical Method: Quantitative

Detection Method: Colorimetric

Specificity: This assay has high sensitivity and excellent specificity for detection of Mini Samples Fibroblast Growth Factor 22 (FGF22).

No significant cross-reactivity or interference between Mini Samples Fibroblast Growth Factor 22 (FGF22) and analogues was observed.

Components:

- Pre-coated, ready to use 96-well strip plate
- Plate sealer for 96 wells
- Standard Diluent
- Assay Diluent A
- Assay Diluent B
- Stop Solution
- Standard
- Detection Reagent A
- Detection Reagent B

Product Details

- TMB Substrate
- Wash Buffer (30 x concentrate)
- Instruction manual

- Material not included:
- Microplate reader with 450 nm filter.
 - Precision single or multi-channel pipettes and disposable tips.
 - Eppendorf Tubes for diluting samples.
 - Deionized or distilled water.
 - Absorbent paper for blotting the microtiter plate.
 - Container for Wash Solution

Target Details

Target: FGF22

Alternative Name: Fibroblast Growth Factor 22 ([FGF22 Products](#))

UniProt: [Q9ESS2](#)

Pathways: [RTK Signaling](#), [Fc-epsilon Receptor Signaling Pathway](#), [EGFR Signaling Pathway](#), [Neurotrophin Signaling Pathway](#)

Application Details

- Application Notes:
- Limited by the current condition and scientific technology, we cannot completely conduct the comprehensive identification and analysis on the raw material provided by suppliers. So there might be some qualitative and technical risks to use the kit.
 - The final experimental results will be closely related to validity of the products, operation skills of the end users and the experimental environments. Please make sure that sufficient samples are available.
 - Kits from different batches may be a little different in detection range, sensitivity and color developing time.
 - Do not mix or substitute reagents from one kit lot to another. Use only the reagents supplied by manufacturer.
 - Protect all reagents from strong light during storage and incubation. All the bottle caps of reagents should be covered tightly to prevent the evaporation and contamination of microorganism.
 - There may be some foggy substance in the wells when the plate is opened at the first time. It will not have any effect on the final assay results. Do not remove microtiter plate from the storage bag until needed.
 - Wrong operations during the reagents preparation and loading, as well as incorrect parameter setting for the plate reader may lead to incorrect results. A microplate plate reader with a bandwidth of 10nm or less and an optical density range of 0-3 O.D. or greater at 450 ±

Application Details

10nm wavelength is acceptable for use in absorbance measurement. Please read the instruction carefully and adjust the instrument prior to the experiment.

- Even the same operator might get different results in two separate experiments. In order to get better reproducible results, the operation of every step in the assay should be controlled. Furthermore, a preliminary experiment before assay for each batch is recommended.
- Each kit has been strictly passed Q.C test. However, results from end users might be inconsistent with our in-house data due to some unexpected transportation conditions or different lab equipments. Intra-assay variance among kits from different batches might arise from above factors, too.
- Kits from different manufacturers for the same item might produce different results, since we have not compared our products with other manufacturers.

Comment:

Information on standard material:

The standard might be recombinant protein or natural protein, that will depend on the specific kit. Moreover, the expression system is E.coli or yeast or mammal cell. There is 0.05% proclin 300 in the standard as preservative.

Information on reagents:

The stop solution used in the kit is sulfuric acid with concentration of 1 mol/L. And the wash solution is TBS. The standard diluent contains 0.02 % sodium azide, assay diluent A and assay diluent B contain 0.01% sodium azide. Some kits can contain is BSA in them.

Information on antibodies:

The provided antibodies and their host vary in different kits.

Sample Volume:

25 μ L

Assay Time:

3 h

Plate:

Pre-coated

Protocol:

The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Mini Samples Fibroblast Growth Factor 22 (FGF22). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Mini Samples Fibroblast Growth Factor 22 (FGF22). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Mini Samples Fibroblast Growth Factor 22 (FGF22), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm \pm 10nm. The concentration of Mini Samples

Application Details

Fibroblast Growth Factor 22 (FGF22) in the samples is then determined by comparing the O.D. of the samples to the standard curve.

- Sample Preparation:
- We are only responsible for the kit itself, but not for the samples consumed during the assay. The user should calculate the possible amount of the samples used in the whole test. Please reserve sufficient samples in advance.
 - Please predict the concentration before assaying. If values for these are not within the range of the standard curve, users must determine the optimal sample dilutions for their particular experiments. Sample should be diluted by 0.01mol/L PBS(PH=7.0-7.2).
 - If the samples are not indicated in the manual, a preliminary experiment to determine the validity of the kit is necessary.
 - Tissue or cell extraction samples prepared by chemical lysis buffer may cause unexpected ELISA results due to the impacts from certain chemicals.
 - Due to the possibility of mismatching between antigen from other origin and antibody used in our kits (e.g. antibody targets conformational epitope rather than linear epitope), some native or recombinant proteins from other manufacturers may not be recognized by our products.
 - Influenced by the factors including cell viability, cell number or sampling time, samples from cell culture supernatant may not be detected by the kit.
 - Fresh samples without long time storage is recommended for the test. Otherwise, protein degradation and denaturalization may occur in those samples and finally lead to wrong results.
-

- Assay Procedure:
1. Prepare all reagents, samples and standards,
 2. Add 100µL standard or sample to each well. Incubate 2 hours at 37 °C,
 3. Aspirate and add 100µL prepared Detection Reagent A. Incubate 1 hour at 37 °C,
 4. Aspirate and wash 3 times,
 5. Add 100µL prepared Detection Reagent B. Incubate 30 minutes at 37 °C,
 6. Aspirate and wash 5 times,
 7. Add 90µL Substrate Solution. Incubate 10-20 minutes at 37 °C,
 8. Add 50µL Stop Solution. Read at 450nm immediately.
-

Assay Precision:

Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Mini Samples Fibroblast Growth Factor 22 (FGF22) were tested 20 times on one plate, respectively.

Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Mini Samples Fibroblast Growth Factor 22 (FGF22) were tested on 3 different plates, 8 replicates in each plate.

$CV(\%) = SD/mean \times 100$

Intra-Assay: CV<10%

Inter-Assay: CV<12%

Restrictions: For Research Use only

Handling

Precaution of Use:	The Stop Solution suggested for use with this kit is an acid solution. Wear eye, hand, face, and clothing protection when using this material.
Handling Advice:	<p>The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition.</p> <p>To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.</p>
Storage:	4 °C
Storage Comment:	<ul style="list-style-type: none">• For unopened kit: All the reagents should be kept according to the labels on vials. The Standard, Detection Reagent A, Detection Reagent B and the 96-well strip plate should be stored at -20°C upon receipt while the others should be at 4°C.• For opened kit: When the kit is opened, the remaining reagents still need to be stored according to the above storage condition. Besides, please return the unused wells to the foil pouch containing the desiccant pack, and reseal along entire edge of zip-seal. Note: It is highly recommended to use the remaining reagents within 1 month provided this is within the expiration date of the kit.• For ELISA kit, 1 day storage at 37°C can be considered as 2 months at 4°C, which means 3 days at 37°C equaling 6 months at 4°C.
Expiry Date:	6 months