

## Datasheet for ABIN4969818

## **HGS ELISA Kit**



| $\sim$ |     |   |    |     |     |
|--------|-----|---|----|-----|-----|
|        | 1// | Р | rv | I P | ۱۸/ |
|        |     |   |    |     |     |

| Overview                 |  |  |
|--------------------------|--|--|
| Quantity:                | 96 tests   |  |
| Target:                  | HGS  |  |
| Reactivity:              | Human  |  |
| Method Type:             | Sandwich ELISA   |  |
| Detection Range:         | 15.625 pg/mL - 1000 pg/mL  |  |
| Minimum Detection Limit: | 15.625 pg/mL   |  |
| Application:             | ELISA  |  |
| Product Details          |  |  |
| Purpose:                 | Human Hepatocyte growth factor-regulated tyrosine kinase substrate ELISA Kit is an ELISA Kit |  |
|                          | against Hepatocyte growth factor-regulated tyrosine kinase substrate.                        |  |
| Sample Type:             | Plasma, Serum  |  |
| Analytical Method:       | Quantitative   |  |
| Detection Method:        | Colorimetric   |  |
| Sensitivity:             | 9.38 pg/mL   |  |
| Target Details           |  |  |
| Target:                  | HGS  |  |
| Alternative Name:        | HGS (HGS Products)   |  |
|                          | EGFR Signaling Pathway, CXCR4-mediated Signaling Events, Synaptic Vesicle Exocytosis, EGFR   |  |

## Downregulation

| Application Details |   |  |  |
|---------------------|---|--|--|
| Application Notes:  | Stability: The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5 % within the expiration date under appropriate storage conditions. To minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same user throughout. Recommended dilutions: Optimal dilutions/concentrations should be determined by the end user.  Standard Form: Lyophilized |  |  |
| Comment:            | The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5% within the expiration date under appropriate storage conditions. To minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same user throughout.   |  |  |
| Plate:              | Pre-coated  |  |  |
| Restrictions:       | For Research Use only   |  |  |
| Handling            |   |  |  |
| Storage:            | 4 °C/-20 °C   |  |  |
| Storage Comment:    | Upon receipt, store the kit according to the storage instruction in the kit's manual.   |  |  |
| Expiry Date:        | 6 months  |  |  |