

# Datasheet for ABIN4971518

## **NPS ELISA Kit**



### Overview

| Quantity:                | 96 tests              |
|--------------------------|-----------------------|
| Target:                  | NPS                   |
| Reactivity:              | Human                 |
| Method Type:             | Sandwich ELISA        |
| Detection Range:         | 78 pg/mL - 5000 pg/mL |
| Minimum Detection Limit: | 78 pg/mL              |
| Application:             | ELISA                 |
|                          |                       |

| Product Details    |  |
|--------------------|--|
| Purpose:           | Human Neuropeptide S ELISA Kit is an ELISA kit against Human Neuropeptide S (NPS). |
| Sample Type:       | Plasma, Serum  |
| Analytical Method: | Quantitative   |
| Detection Method:  | Colorimetric   |
| Target Details     |  |
| Target:            | NPS  |
| Abstract:          | NPS Products   |

# **Application Details**

Stability: The stability of the kit is determined by the rate of activity loss. The loss rate is less Application Notes:

# **Application Details**

|                  | than 5 % within the expiration date under appropriate storage conditions. To minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same user throughout. Recommended dilutions: Optimal dilutions/concentrations should be determined by the end user. |
|------------------|---|
| Comment:         | The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5% within the expiration date under appropriate storage conditions. To minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same user throughout.     |
| Plate:           | Pre-coated  |
| Restrictions:    | For Research Use only   |
| Handling         |   |
| Storage:         | 4 °C/-20 °C   |
| Storage Comment: | Upon receipt, store the kit according to the storage instruction in the kit's manual.   |
| Expiry Date:     | 6 months  |