

Datasheet for ABIN4973662

Urocortin ELISA Kit



Overview

Quantity:	96 tests
Target:	Urocortin (UCN)
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	15.6 pg/mL - 1000 pg/mL
Minimum Detection Limit:	15.6 pg/mL
Application:	ELISA

Product Details

Purpose:	Human Urocortin ELISA Kit is an ELISA kit against Human Urocortin (UCN).
Sample Type:	Plasma, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric

Target Details

Target:	Urocortin (UCN)
Alternative Name:	Urocortin (UCN Products)
Pathways:	Positive Regulation of Peptide Hormone Secretion, Hormone Activity, Negative Regulation of

Hormone Secretion, cAMP Metabolic Process, Regulation of Cell Size, Feeding Behaviour

Application Details

Application Notes:	Stability: The stability of the kit is determined by the rate of activity loss. The loss rate is less
	than 5 % within the expiration date under appropriate storage conditions. To minimize
	performance fluctuations, operation procedures and lab conditions should be strictly controlled.
	It is also strongly suggested that the whole assay is performed by the same user throughout.
	Recommended dilutions: Optimal dilutions/concentrations should be determined by the end
	user.
Comment:	The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5%
	within the expiration date under appropriate storage conditions. To minimize performance
	fluctuations, operation procedures and lab conditions should be strictly controlled. It is also
	strongly suggested that the whole assay is performed by the same user throughout.
Plate:	Pre-coated
Restrictions:	For Research Use only
Handling	
Storage:	4 °C/-20 °C
Storage Comment:	Upon receipt, store the kit according to the storage instruction in the kit's manual.
Expiry Date:	6 months