antibodies - online.com







CDK4 ELISA Kit





_						
0	V	е.	r٧	1	6	W

Quantity:	96 tests
Target:	CDK4
Reactivity:	Mouse
Method Type:	Sandwich ELISA
Detection Range:	62.5-4000 pg/mL
Minimum Detection Limit:	62.5 pg/mL
Application:	ELISA
Product Details	
Purpose:	The kit is a sandwich enzyme immunoassay for in vitro quantitative measurement of CDK4 in
	Mouse serum, plasma and other biological fluids
Sample Type:	Plasma, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This kit recognizes natural and recombinantMouseCDK4. No significant cross-reactivity or
	interference between MouseCDK4 and analogues was observed. Note: Limited by existing
	techniques, cross reaction may still exist, as it is impossible for us to complete the cross-
	reactivity detection between MouseCDK4 and all the analogues.
Sensitivity:	37.5 pg/mL
Components:	Micro ELISA Plate
	Reference Standard

Concentrated Biotinylated Detection Ab Concentrated HRP Conjugate Reference Standard & Sample Diluent Biotinylated Detection Ab Diluent HRP Conjugate Diluent Concentrated Wash Buffer (25×) Substrate Reagent Stop Solutio Plate Sealer **Product Description** Certificate of Analysis Material not included: Microplate reader with 450nm wavelength filter High-precision transferpettor, EP tubes and disposable pipette tips 37°C Incubator Deionized or distilled water Absorbent paper Loading slot for Wash Buffer **Target Details**

Target:	CDK4	
Alternative Name:	Cyclin Dependent Kinase 4 (CDK4) (CDK4 Products)	
Pathways:	Cell Division Cycle, Mitotic G1-G1/S Phases, Regulation of Cell Size	
Application Details		
Comment:	Information on standard material:	
	The formulation of the standard is 0.01 M PBS. The standard contains additives (1 % BSA).	
	Information on reagents:	
	Reagents include 1 M $\rm SO_2$. Azide, thimerosal, 2-mercaptoethanol (2-ME) or any other poisonous	
	materials are not used.	
	Information on antibodies:	
	The provided antibodies and their host vary in different kits. All antibodies are affinity purified	

Application Details

The sensitivity of this assay, or Lower Limit of Detection (LLD) was defined as the lowest protein concentration that could be differentiated from zero. It was determined by adding two standard deviations to the mean optical density value of twenty zero standard replicates and calculating the corresponding concentration.

Sample Volume:

100 μL

Plate:

Pre-coated

Protocol:

This ELISA kit uses Sandwich-ELISA as the method. The micro ELISA plate provided in this kit has been pre-coated with an antibody specific to CDK4. Standards or samples are added to the appropriate micro ELISA plate wells and combined with the specific antibody. Then a biotinylated detection antibody specific for CDK4and Avidin-Horseradish Peroxidase (HRP) conjugate is added to each micro plate well successively and incubated. Free components are washed away. The substrate solution is added to each well. Only those wells that contain CDK4, biotinylated detection antibody and Avidin-HRP conjugate will appear blue in color. The enzyme-substrate reaction is terminated by the addition of a sulphuric acid solution and the color turns yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 nm \pm 2 nm. The OD value is proportional to the concentration of CDK4. You can calculate the concentration of CDK4in the samples by comparing the OD of the samples to the standard curve. 6th Edition, revised in June, 2015 www.elabscience.com 4

Restrictions:

For Research Use only

Handling

Handling Advice:


All the reagents in the kit should be stored according to the labels on vials. Unused wells should be returned to the foil pouch with the desiccant pack and resealed along entire edge of zip-seal. Substrate Reagent shouldn't be kept at -20 °C (Check!). Exposure of reagents to strong light should be avoided in the process of incubation and storage. All the taps of reagents should be tightened to prevent evaporation and microbial contamination. If not to store reagents according to above suggestions, erroneous results may occur.

Storage:

4 °C/-20 °C

Storage Comment:

The unopened kit can be stored at 4°C for 1 month. If the kit is not used within 1 month, store the items separately according to the conditions since the kit is received.



ELISA

Image 1. Typical standard curve