antibodies - online.com







PPARGC1B ELISA Kit





Co to i roddot page

\sim					
()	VE	۲۱	/1	\triangle	Λ


Quantity:	96 tests
Target:	PPARGC1B
Reactivity:	Mouse
Method Type:	Sandwich ELISA
Detection Range:	0.156-10 ng/mL
Minimum Detection Limit:	0.156 ng/mL
Application:	ELISA
Product Details	
Purpose:	The kit is a sandwich enzyme immunoassay for in vitro quantitative measurement of PPARγC1
	β in Mouse serum, plasma and other biological fluids
Sample Type:	Plasma, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This kit recognizes natural and recombinantMousePPARyC1β. No significant cross-reactivity or
	interference between MousePPAR γ C1 β and analogues was observed. Note: Limited by existing
	techniques, cross reaction may still exist, as it is impossible for us to complete the cross-
	reactivity detection between MousePPARγC1β and all the analogues.
Sensitivity:	0.094 ng/mL
Components:	Micro ELISA Plate
	Reference Standard

	Concentrated Biotinylated Detection Ab
	Concentrated HRP Conjugate
	Reference Standard & Sample Diluent
	Biotinylated Detection Ab Diluent
	HRP Conjugate Diluent
	Concentrated Wash Buffer (25×)
	Substrate Reagent
	Stop Solutio
	Plate Sealer
	Product Description
	Certificate of Analysis
Material not included:	Microplate reader with 450nm wavelength filter
	High-precision transferpettor, EP tubes and disposable pipette tips
	37 °C Incubator
	Deionized or distilled water
	Absorbent paper
	Loading slot for Wash Buffer
Target Details	
Target:	PPARGC1B
Alternative Name:	Peroxisome Proliferator Activated Receptor Gamma Coactivator 1 beta (PPARgammaC1beta) (
	PPARGC1B Products)
Pathways:	AMPK Signaling, Intracellular Steroid Hormone Receptor Signaling Pathway, Regulation of Lipid
	Metabolism by PPARalpha
Application Details	
Application Details	
Comment:	Information on standard material:
Comment:	Information on standard material: The formulation of the standard is 0.01 M PBS. The standard contains additives (1 % BSA).
Comment:	
Comment:	The formulation of the standard is 0.01 M PBS. The standard contains additives (1 % BSA).

Information on antibodies:

Application Details	
	The provided antibodies and their host vary in different kits. All antibodies are affinity purified
	The sensitivity of this assay, or Lower Limit of Detection (LLD) was defined as the lowest
	protein concentration that could be differentiated from zero. It was determined by adding two
	standard deviations to the mean optical density value of twenty zero standard replicates and
	calculating the corresponding concentration.
Sample Volume:	100 μL
Plate:	Pre-coated
Protocol:	This ELISA kit uses Sandwich-ELISA as the method. The micro ELISA plate provided in this kit
	has been pre-coated with an antibody specific to PPAR _Y C1 _B . Standards or samples are added
	to the appropriate micro ELISA plate wells and combined with the specific antibody. Then a
	biotinylated detection antibody specific for PPARγC1βand Avidin-Horseradish Peroxidase (HRP
	conjugate is added to each micro plate well successively and incubated. Free components are
	washed away. The substrate solution is added to each well. Only those wells that contain PPAR
	$\gamma C1\beta$, biotinylated detection antibody and Avidin-HRP conjugate will appear blue in color. The
	enzyme-substrate reaction is terminated by the addition of a sulphuric acid solution and the
	color turns yellow. The optical density (OD) is measured spectrophotometrically at a
	wavelength of 450 nm \pm 2 nm. The OD value is proportional to the concentration of PPAR γ C1 β .
	You can calculate the concentration of PPAR γ C1 β in the samples by comparing the OD of the
	samples to the standard curve. 6th Edition, revised in June, 2015 www.elabscience.com 4
Restrictions:	For Research Use only
Handling	
Handling Advice:	All the reagents in the kit should be stored according to the labels on vials. Unused wells should
	be returned to the foil pouch with the desiccant pack and resealed along entire edge of zip-seal.
	Substrate Reagent shouldn't be kept at -20 °C (Check!). Exposure of reagents to strong light
	should be avoided in the process of incubation and storage. All the taps of reagents should be
	tightened to prevent evaporation and microbial contamination. If not to store reagents
	according to above suggestions, erroneous results may occur.
Storage:	4 °C/-20 °C
Storage Comment:	The unopened kit can be stored at 4°C for 1 month. If the kit is not used within 1 month,

store the items separately according to the conditions since the kit is received.



ELISA

Image 1. Typical standard curve