



[Go to Product page](#)

Datasheet for ABIN5067552

Cholic Acid ELISA Kit

3 Images

Overview

Quantity:	96 tests
Target:	Cholic Acid (CA)
Reactivity:	Chemical
Method Type:	Competition ELISA
Application:	ELISA

Product Details

Purpose:	The Cholic Acid ELISA kit is a competitive ELISA for the quantitative measurement of cholic acid.
Sample Type:	Cell Samples, Plasma, Serum, Tissue Samples
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Sensitivity:	0.4 μ M
Characteristics:	The Cholic Acid ELISA Kit is a competitive enzyme immunoassay developed for rapid detection and quantitation of cholic acid in serum, feces, or other cell or tissue samples. The quantity of cholic acid in unknown samples is determined by comparing its absorbance with that of a known cholic acid standard curve. The kit has detection sensitivity limit of 0.4 μ M cholic acid. Each Cholic Acid ELISA Kit provides sufficient reagents to perform up to 96 assays, including standard curve and unknown samples.
Components:	<ol style="list-style-type: none">96-well Protein Binding Plate : One strip well 96-well plate.Anti-Cholic Acid Antibody (500X) : One 10 μL vial.

Product Details

3. Secondary Antibody, HRP Conjugate : One 20 µL vial.
4. Assay Diluent : One 50 mL bottle.
5. 10X Wash Buffer : One 100 mL bottle.
6. Substrate Solution : One 12 mL amber bottle.
7. Stop Solution (Part. No. 310808): One 12 mL bottle.
8. Cholic Acid Standard : One 100 µL vial of 2.5 mM Cholic Acid in 50 mM KH₂PO₄, pH 7.4, 25 mM NaCl.

Box 2 (shipped on blue ice packs)

Target Details

Target: Cholic Acid (CA)

Alternative Name: Cholic Acid ([CA Products](#))

Target Type: Chemical

Background: Cholic acid is a primary bile acid. Along with chenodeoxycholic acid, cholic acid is one of the two major bile acids synthesized from cholesterol by the liver. Bile is a complex mixture of lipids, protein, carbohydrates, mineral salts, vitamins, and various trace elements, with bile acids making up about 67 % of the total composition. Bile acids are produced from excess cholesterol, secreted from the liver, absorbed into the small intestines, and returned to the liver with portal blood. While bile acid synthesis is critical for the removal of cholesterol from the body, bile acids are also needed for proper uptake of dietary lipids, fat soluble vitamins, and other nutrients into the small intestines. Under physiological conditions, newly synthesized bile acids are conjugated to glycine or taurine to form bile salts, and not much free bile acid is actually found in bile. Determining circulatory levels of bile acids can be used to identify or diagnose certain liver diseases. In addition, elevated serum bile levels have been observed in intrahepatic cholestasis of pregnancy cases. Recently, bile acids have been recognized as integrators of metabolic processes as well as signaling molecules.

Application Details

Application Notes: Optimal working dilution should be determined by the investigator.

Comment:

- Detection sensitivity limit of 0.4µM cholic acid
- Suitable for use with plasma, serum, urine, feces, and cell and tissue lysates
- Cholic acid standard included

Plate: Uncoated

Protocol: The unknown cholic acid samples or cholic acid standards are first added to a Cholic Acid

Application Details

Conjugate preadsorbed microplate. After a brief incubation, an Anti-Cholic Acid monoclonal antibody is added, followed by an HRP conjugated secondary antibody. The cholic acid content in unknown samples is determined by comparison with predetermined cholic acid standard curve.

Reagent Preparation:

Cholic Acid Conjugate Coated Plate: Dilute the proper amount of 100X Cholic Acid Conjugate 1:100 into 1X PBS. Add 100 μ L of the diluted 1X Cholic Acid Conjugate to each well and incubate at 37 °C for two hours or overnight at 4 °C. Remove the coating solution and wash twice with 200 L of 1X PBS. Blot plate on paper towels to remove excess fluid. Add 200 μ L of Assay Diluent to each well and block for 1 hr at room temperature. Transfer the plate to 4 °C and remove the Assay Diluent immediately before use. Note: The Cholic Acid-Conjugate coated wells are not stable and should be used within 24 hrs after coating. Only coat the number of wells to be used immediately. **1X Wash Buffer:** Dilute the 10X Wash Buffer to 1X with deionized water. Stir to homogeneity. **Anti-Cholic Acid Antibody and Secondary Antibody:** Immediately before use dilute the Anti- Cholic Acid Antibody 1:500 and Secondary Antibody 1:1000 with Assay Diluent. Do not store diluted solutions. **Preparation of Standard Curve** 1. Prepare Cholic Acid standards by diluting in 1X Assay Diluent. First, dilute the stock Cholic Acid Standard 2.5 mM solution 1:10 in 1X Assay Diluent for a 250 μ M solution. (e.g. add 10 μ L of the stock 2.5 mM standard to 90 μ L of 1X Assay Diluent). Vortex thoroughly. 2. Use this 250 μ M solution to prepare a series of the remaining cholesterol standards according to Table 1 below.

250 μ M Cholic Acid Assay Diluent	Cholic Acid Standard	Tubes	Standard (μ L)	(μ L)	(M)
1	10	990	25	2	500
of Tube #1	500	12.5	3	500	
of Tube #2	500	6.25	4	500	
of Tube #3	500	3.13	5	500	
of Tube #4	500	1.56	6	500	
of Tube #5	500	0.78	7	500	
of Tube #6	500	0.39	8	0	500
0	500	0			

Table 1. Preparation of Cholic Acid Standards.

Sample Preparation:

Serum: Avoid hemolyzed and lipemic blood samples. Collect blood in a tube with no anticoagulant. Allow the blood to clot at room temperature for 30 minutes. Centrifuge at 2500 x g for 20 minutes. Remove the yellow serum supernatant without disturbing the white buffy layer. Aliquot samples for testing and store at -80 °C. Perform dilutions in Assay Diluent as necessary. **4 Plasma:** Avoid hemolyzed and lipemic blood samples. Collect blood with heparin or citrate and centrifuge at 2000 x g and 4 °C for 10 minutes. Remove the plasma layer and store on ice. Avoid disturbing the white buffy layer. Aliquot samples for testing and store at -80 °C. Perform dilutions in Assay Diluent as necessary. **Cells, tissues, or feces:** Homogenize 50-200 mg of the cell pellet, tissue, or feces in 0.5-2 mL of ice cold PBS using a mortar and pestle or by dounce homogenization. Incubate the homogenate at 4 °C for 20 minutes. Transfer the homogenate to a centrifuge tube and centrifuge at 12000 x g for 20 minutes. Recover the supernatant and transfer to a fresh tube. Store resuspended sample at -20 °C or colder until

Application Details


ready to test by ELISA. Perform dilutions in Assay Diluent as necessary.

Assay Procedure:	<ol style="list-style-type: none">1. Prepare and mix all reagents thoroughly before use. Each cholic acid sample including unknown and standard should be assayed in duplicate.2. Add 50 μL of unknown sample or Cholic Acid standards to the wells of the Cholic Acid Conjugate coated plate. Incubate at room temperature for 10 minutes on an orbital shaker.3. Add 50 μL of the diluted Anti-Cholic Acid antibody to each well, incubate at room temperature for 1 hour on an orbital shaker.4. Wash microwell strips 3 times with 250 μL 1X Wash Buffer per well with thorough aspiration between each wash. After the last wash, empty wells and tap microwell strips on absorbent pad or paper towel to remove excess 1X Wash Buffer.5. Add 100 μL of the diluted Secondary Antibody-HRP Enzyme Conjugate to all wells.6. Incubate at room temperature for 1 hour on an orbital shaker.7. Wash microwell strips 3 times according to step 4 above. Proceed immediately to the next step.8. Warm Substrate Solution to room temperature. Add 100 μL of Substrate Solution to each well, including the blank wells. Incubate at room temperature on an orbital shaker. Actual incubation time may vary from 2-30 minutes. Note: Watch plate carefully, if color changes rapidly, the reaction may need to be stopped sooner to prevent saturation.9. Stop the enzyme reaction by adding 100 μL of Stop Solution into each well, including the blank wells. Results should be read immediately (color will fade over time).10. Read absorbance of each microwell on a spectrophotometer using 450 nm as the primary wave length. 5
------------------	---

Restrictions:	For Research Use only
---------------	-----------------------


Handling

Handling Advice:	Avoid multiple freeze/thaw cycles.
Storage:	4 °C/-20 °C
Storage Comment:	Upon receipt, aliquot and store both the Anti-Cholic Acid Antibody and 100X Cholic Acid Conjugate at -20°C and avoid multiple freeze/thaw cycles. Store all other components at 4°C. 3




Biochemical Assay

Image 1. Cholic Acid Levels in undiluted Human or Rat Serum compared to Negative Control (Assay Diluent)



Biochemical Assay

Image 2. Cholic Acid ELISA Standard Curve



Biochemical Assay

Image 3. Cholic Acid Levels in Liver. Chicken liver was homogenized, and the homogenate supernatant (11.4 mg/mL protein by BCA assay) was diluted in Assay Diluent and analyzed according to the Assay Protocol.