antibodies -online.com







ULK1 ELISA Kit





()	1/0	r\ /1	014	
()	ve	I V I	-v	V

96 tests
ULK1
pSer556, total
Human, Mouse
Sandwich ELISA
ELISA
Human and Mouse Phospho-ULK1 (S556) and Total ULK1 ELISA Kit. This assay semi- quantitatively measures ULK1 phosphorylated at Serine-556 as well as total ULK1 in cell lysate samples.
Cell Culture Lysate
Semi-Quantitative
Colorimetric
This ELISA kit recognizes Human and Mouse ULK1 phosphorylated at site Serine-556 as well as total ULK1.
 Pre-Coated 96-well Strip Microplate Wash Buffer Anti-Phospho Antibody Anti-Pan Antibody HRP-Conjugated Secondary Antibody Streptavidin-Conjugated HRP

Product Details

- · Assay Diluent
- · TMB One-Step Substrate
- · Stop Solution
- · Lysis Buffer
- · Positive Control Sample

Components:

- · Pre-Coated 96-well Strip Microplate
- · Wash Buffer
- · Anti-Phospho Antibody
- Anti-Pan Antibody
- · HRP-Conjugated Secondary Antibody
- · Streptavidin-Conjugated HRP
- · Assay Diluent
- TMB One-Step Substrate
- · Stop Solution
- · Lysis Buffer
- · Positive Control Sample

Material not included:

- · Distilled or deionized water
- · 100 mL and 1 liter graduated cylinders
- · Tubes to prepare sample dilutions
- · Protease and Phosphatase inhibitors
- Precision pipettes to deliver 2 µL to 1 mL volumes
- · Adjustable 1-25 mL pipettes for reagent preparation
- · Benchtop rocker or shaker
- Microplate reader capable of measuring absorbance at 450 nm

Target Details

Target:	ULK1
Alternative Name:	ULK1 (ULK1 Products)
Gene ID:	8408
UniProt:	075385, 070405
Pathways:	Regulation of Cell Size, Autophagy

Application Details

Application Notes:	Optimal working dilution should be determined by the investigator.	
Plate:	Pre-coated	

Application Details

Protocol:

- 1. Prepare all reagents and samples as instructed in the manual.
- 2. Add 100 µL of sample or positive control to each well.
- 3. Incubate 2.5 h at RT or O/N at 4 °C.
- 4. Add 100 μL of prepared primary antibody to each well.
- 5. Incubate 1 h at RT.
- 6. Add 100 µL of prepared 1X HRP-Streptavidin to each well.
- 7. Incubate 1 h at RT.
- 8. Add 100 µL of TMB One-Step Substrate Reagent to each well.
- 9. Incubate 30 min at RT.
- 10. Add 50 µL of Stop Solution to each well.
- 11. Read at 450 nm immediately.

Restrictions:

For Research Use only

Handling

Storage:

-20 °C


Storage Comment:

Upon receipt, the kit should be stored at -20 °C. Please use within 6 months from the date of shipment. After initial use, Wash Buffer Concentrate (Item B), Assay Diluent (Item E), TMB One-Step Substrate Reagent (Item H), HRP-Streptavidin (Item G), Stop Solution (Item I) and Cell Lysate Buffer (Item J) should be stored at 4 °C to avoid repeated freeze-thaw cycles. Return unused wells to the pouch containing desiccant pack, reseal along entire edge and store at -20 °C. Reconstituted Positive Control (Item K) should be stored at -70 °C.

Expiry Date:

6 months

Images



ELISA

Image 1. C2C12 cells were untreated or treated with 1mM H2O2 for 15 min. Cell lysates were analyzed using this phosphoELISA and Western Blot.





Image 2. C2C12 cells were untreated or treated with 1mM

H2O2 for 15 min. Cell lysates were analyzed using this phosphoELISA and Western Blot.



ELISA

Image 3. Jurkat cells were treated with H2O2 at 37° C for 2 min. Cells were solubilzed at $4 \times 10^{\circ}$ 7 cells/ml in Cell Lysate Buffer. Serial dilutions of lysates were analyzed in this ELISA.