antibodies .- online.com





Datasheet for ABIN5526736

ATR ELISA Kit

3 Ima

Images



Overview

96 tests
ATR
pThr1989
Human
Sandwich ELISA
ELISA
Human Phospho-ATR (Thr1989) ELISA Kit. This assay semi-quantitatively measures ATR phosphorylated at Threonnine-1989 in cell lysate samples.
Cell Culture Lysate
Semi-Quantitative
Colorimetric
This ELISA kit recognizes Human ATR phosphorylated at site Threonnine-1989.
 Pre-Coated 96-well Strip Microplate Wash Buffer Anti-Phospho Antibody HRP-Conjugated Secondary Antibody Assay Diluent TMB One-Step Substrate Stop Solution

Product Details

	Positive Control Sample
Components:	Pre-Coated 96-well Strip Microplate
	Wash Buffer
	Anti-Phospho Antibody
	HRP-Conjugated Secondary Antibody
	Assay Diluent
	TMB One-Step Substrate
	Stop Solution
	Lysis Buffer
	Positive Control Sample
Material not included:	Distilled or deionized water
	100 mL and 1 liter graduated cylinders
	Tubes to prepare sample dilutions
	Protease and Phosphatase inhibitors
	 Precision pipettes to deliver 2 μL to 1 mL volumes
	Adjustable 1-25 mL pipettes for reagent preparation
	Benchtop rocker or shaker
	 Microplate reader capable of measuring absorbance at 450 nm

Target Details

Target:	ATR
Alternative Name:	ATR (ATR Products)
Gene ID:	545
UniProt:	Q13535
Pathways:	Positive Regulation of Response to DNA Damage Stimulus

Application Details

Application Notes:	Optimal working dilution should be determined by the investigator.
Plate:	Pre-coated
Protocol:	 Prepare all reagents and samples as instructed in the manual. Add 100 μL of sample or positive control to each well. Incubate 2.5 h at RT or O/N at 4 °C. Add 100 μL of prepared primary antibody to each well. Incubate 1 h at RT.

- 6. Add 100 µL of prepared 1X HRP-Streptavidin to each well.
- 7. Incubate 1 h at RT.
- 8. Add 100 µL of TMB One-Step Substrate Reagent to each well.
- 9. Incubate 30 min at RT.
- 10. Add 50 µL of Stop Solution to each well.
- 11. Read at 450 nm immediately.

Restrictions:

For Research Use only

Handling

Storage: -20 °C


Storage Comment:


Upon receipt, the kit should be stored at -20 °C. Please use within 6 months from the date of shipment. After initial use, Wash Buffer Concentrate (Item B), Assay Diluent (Item E), TMB One-Step Substrate Reagent (Item H), HRP-Streptavidin (Item G), Stop Solution (Item I) and Cell Lysate Buffer (Item J) should be stored at 4 °C to avoid repeated freeze-thaw cycles. Return unused wells to the pouch containing desiccant pack, reseal along entire edge and store at -20 °C. Reconstituted Positive Control (Item K) should be stored at -70 °C.

Expiry Date:

6 months

Images








Anti pan ATR


ELISA

Image 1. T47D cells were untreated or treated with UV. Cell lysates were analyzed using this phosphoELISA and Western Blot.



ELISA

Image 2. T47D cells were exposed to 50J/m2 of UV light followed by a 4 hours recovery period. Cells were solubilzed at 4×10^{4} 7 cells/ml in Cell Lysate Buffer. Serial dilutions of lysates were analyzed in this ELISA.



ELISA

Image 3. T47D cells were untreated or treated with UV. Cell lysates were analyzed using this phosphoELISA and Western Blot.