



[Go to Product page](#)

## Datasheet for ABIN5526752

### **RICTOR ELISA Kit**

#### Overview

Quantity:	96 tests
Target:	RICTOR
Binding Specificity:	pThr1135, total
Reactivity:	Human, Mouse
Method Type:	Cell ELISA
Application:	ELISA

#### Product Details

Purpose:	Cell-Based ELISA Kit. This assay semi-quantitatively measures Rictor phosphorylated at Threonine-1135 as well as total Rictor in adherent cell lines.
Sample Type:	Adherent Cell Culture
Analytical Method:	Semi-Quantitative
Detection Method:	Colorimetric
Specificity:	This ELISA kit recognizes Human and Mouse Rictor phosphorylated at site Threonine-1135 as well as total Rictor.
Characteristics:	<ul style="list-style-type: none"> <li>• Rapidly measure phosphorylated protein in adherent cell lines</li> <li>• Simultaneously measure Phosphorylated protein and pan protein in one experiment (for normalization purpose)</li> <li>• No sample lysis is needed</li> <li>• Compatible with a standard ELISA plate reader</li> </ul>
Components:	<ul style="list-style-type: none"> <li>• Uncoated 96-well Strip Microplate</li> </ul>

## Product Details

- Wash Buffers
- Fixing Solution
- Quenching Buffer
- Blocking Buffer
- Anti-phospho antibody
- Anti-pan antibody
- HRP-Conjugated Secondary Antibody
- TMB One-Step Substrate
- Stop Solution

### Material not included:

- Distilled or deionized water
- 100 mL and 1 liter graduated cylinders
- Tubes to prepare sample dilutions
- Protease and Phosphatase inhibitors
- Precision pipettes to deliver 2 µL to 1 mL volumes
- Adjustable 1-25 mL pipettes for reagent preparation
- Benchtop rocker or shaker
- Microplate reader capable of measuring absorbance at 450 nm

## Target Details

Target: RICTOR

Alternative Name: Rictor ([RICTOR Products](#))

Gene ID: 253260

UniProt: [Q6R327](#), [Q6QI06](#)

Pathways: [Fc-epsilon Receptor Signaling Pathway](#), [EGFR Signaling Pathway](#), [Neurotrophin Signaling Pathway](#), [Regulation of Actin Filament Polymerization](#), [CXCR4-mediated Signaling Events](#)

## Application Details

Application Notes: Optimal working dilution should be determined by the investigator.

Plate: Pre-coated

Protocol:

1. Prepare all reagents and samples as instructed in the manual.
2. Add 100 µL of sample or positive control to each well.
3. Incubate 2.5 h at RT or O/N at 4 °C.
4. Add 100 µL of prepared primary antibody to each well.
5. Incubate 1 h at RT.
6. Add 100 µL of prepared 1X HRP-Streptavidin to each well.

## Application Details

---

7. Incubate 1 h at RT.
8. Add 100 µL of TMB One-Step Substrate Reagent to each well.
9. Incubate 30 min at RT.
10. Add 50 µL of Stop Solution to each well.
11. Read at 450 nm immediately.

---

Restrictions:	For Research Use only
---------------	-----------------------

## Handling

---

Storage:	-20 °C
----------	--------

Storage Comment:	Upon receipt, the kit should be stored at -20 °C. Please use within 6 months from the date of shipment. After initial use, Wash Buffer Concentrate (Item B), Assay Diluent (Item E), TMB One-Step Substrate Reagent (Item H), HRP-Streptavidin (Item G), Stop Solution (Item I) and Cell Lysate Buffer (Item J) should be stored at 4 °C to avoid repeated freeze-thaw cycles. Return unused wells to the pouch containing desiccant pack, reseal along entire edge and store at -20 °C. Reconstituted Positive Control (Item K) should be stored at -70 °C.
------------------	--

---

Expiry Date:	6 months
--------------	----------