antibodies - online.com







S100A16 ELISA Kit





Overview

| Quantity: | 96 tests |
|--------------|----------------|
| Target: | S100A16 |
| Reactivity: | Human |
| Method Type: | Sandwich ELISA |
| Application: | ELISA |

Product Details

| Brand: | AssayMax™ |
|------------------------|--|
| Sample Type: | Cell Culture Cells, Plasma, Serum |
| Analytical Method: | Quantitative |
| Detection Method: | Colorimetric |
| Material not included: | Microplate reader capable of measuring absorbance at 405 nm. Pipettes (1-20 μ L, 20-200 μ L, and multiple channel). Deionized or distilled reagent grade water Incubator (37 °C) |

Target Details

| Target: | S100A16 |
|-------------------|-----------------------------|
| Alternative Name: | S100-A16 (S100A16 Products) |
| Gene ID: | 140576 |
| UniProt: | Q96FQ6 |
| Pathways: | S100 Proteins |

Application Details

| Plate: | Pre-coated |
|-------------------------|---|
| Calculation of Results: | Calculate the mean value of the duplicate or triplicate readings for each standard and sample. To generate a standard curve, plot the graph using the standard concentrations on the x-axis and the corresponding mean 450 nm absorbance (OD) on the y-axis. The best-fit line can be determined by regression analysis using log-log or four-parameter logistic curve-fit. Determine the unknown sample concentration from the standard curve and multiply the value by the dilution factor. |
| Restrictions: | For Research Use only |
| Handling | |
| Handling Advice: | This product is for Research Use Only and is Not For Use In Diagnostic Procedures. Prepare all |
| | reagents (working diluent buffer, wash buffer, standard, biotinylated antibody, and SP |
| | Conjugate) as instructed, prior to running the assay. Prepare all samples prior to running the |
| | assay. The dilution factors for the samples are suggested in this insert. However, the user |
| | should determine the optimal dilution factor. Spin down the SP Conjugate vial and the |
| | biotinylated antibody vial before opening and using contents. The Stop Solution is an acidic |
| | solution. The kit should not be used beyond the expiration date. 2 |
| Storage: | 4 °C,-20 °C |
| Images | |

images




Image 1.