# antibodies .- online.com







# **HSD17B1 ELISA Kit**



**Image** 



#### Overview

Quantity:	96 tests
Target:	HSD17B1
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.781-50 ng/mL
Minimum Detection Limit:	0.781 ng/mL
Application:	ELISA

#### **Product Details**

Π	
Pu	rpose:

The AssayMaxTM Human HSD17B1 ELISA (Enzyme-Linked Immunosorbent Assay) kit is designed for detection of HSD17B1 in human plasma, serum, and cell culture samples. This assay employs a quantitative sandwich enzyme immunoassay technique that measures human HSD17B1 in approximately 4 hours. A polyclonal antibody specific for human HSD17B1 has been pre-coated onto a 96-well microplate with removable strips. HSD17B1 in standards and samples is sandwiched by the immobilized antibody and a biotinylated polyclonal antibody specific for human HSD17B1, which is recognized by a streptavidin-peroxidase (SP) conjugate. All unbound material is washed away and a peroxidase enzyme substrate is added. The color development is stopped and the intensity of the color is measured.

Brand:	AssayMax™
Sample Type:	Cell Culture Cells, Plasma, Serum
Analytical Method:	Quantitative

## **Product Details**

Detection Method:	Colorimetric
Components:	Human HSD17B1 Microplate: A 96-well polystyrene microplate (12 strips of 8 wells) coated
	with a polyclonal antibody against human HSD17B1. Sealing Tapes: Each kit contains 3 precut,
	pressure sensitive sealing tapes that can be cut to fit the format of the individual assay. Human
	HSD17B1 Standard: Human HSD17B1 in a buffered protein base (50 ng, lyophilized, 2 vials).
	Biotinylated Human HSD17B1 Antibody (50x): A 50-fold concentrated biotinylated polyclonal
	antibody against human HSD17B1 (120 l). EIA Diluent Concentrate (10x): A 10-fold
	concentrated buffered protein base (20 ml). Standard Diluent (1x): A buffered protein base with
	stabilizer (2 ml). Wash Buffer Concentrate (20x): A 20-fold concentrated buffered surfactant (30
	ml, 2 bottles). SP Conjugate (100x): A 100-fold concentrate (80 l). Chromogen Substrate (1x): A
	stabilized peroxidase chromogen substrate tetramethylbenzidine (8 ml). Stop Solution (1x): A
	0.5 N hydrochloric acid solution to stop the chromogen substrate reaction (12 ml).
Material not included:	Microplate reader capable of measuring absorbance at 405 nm. Pipettes (1-20 μL, 20-200 μL,
	and multiple channel). Deionized or distilled reagent grade water Incubator (37 °C)
Target Details	
Target:	HSD17B1
Alternative Name:	HSD17B1 (HSD17B1 Products)
Background:	
Background:	Estradiol 17-beta-dehydrogenase 1 (HSD17B1), also known as 17-beta- hydroxysteroid
Background:	Estradiol 17-beta-dehydrogenase 1 (HSD17B1), also known as 17-beta- hydroxysteroid dehydrogenase type 1, is a member of 17-beta-hydroxysteroid dehydrogenases that compose a
Background:	
Background:	dehydrogenase type 1, is a member of 17-beta-hydroxysteroid dehydrogenases that compose a
Background:	dehydrogenase type 1, is a member of 17-beta-hydroxysteroid dehydrogenases that compose a family of 14 enzymes and catalyze the conversion between the low active 17-ketosteroids and
Background:	dehydrogenase type 1, is a member of 17-beta-hydroxysteroid dehydrogenases that compose a family of 14 enzymes and catalyze the conversion between the low active 17-ketosteroids and the highly active 17 beta-hydroxysteroids. Human HSD17B1 contains 327 amino acids and
Background:	dehydrogenase type 1, is a member of 17-beta-hydroxysteroid dehydrogenases that compose a family of 14 enzymes and catalyze the conversion between the low active 17-ketosteroids and the highly active 17 beta-hydroxysteroids. Human HSD17B1 contains 327 amino acids and exists as a homodimer with two identical subunits of 34.5 kDa (1-2). It uses NADPH as a co-
Background:	dehydrogenase type 1, is a member of 17-beta-hydroxysteroid dehydrogenases that compose a family of 14 enzymes and catalyze the conversion between the low active 17-ketosteroids and the highly active 17 beta-hydroxysteroids. Human HSD17B1 contains 327 amino acids and exists as a homodimer with two identical subunits of 34.5 kDa (1-2). It uses NADPH as a cofactor to predominantly catalyze the reduction of inactive estrone (E1) to the active estrogen 17
Background:	dehydrogenase type 1, is a member of 17-beta-hydroxysteroid dehydrogenases that compose a family of 14 enzymes and catalyze the conversion between the low active 17-ketosteroids and the highly active 17 beta-hydroxysteroids. Human HSD17B1 contains 327 amino acids and exists as a homodimer with two identical subunits of 34.5 kDa (1-2). It uses NADPH as a cofactor to predominantly catalyze the reduction of inactive estrone (E1) to the active estrogen 17 beta-estradiol (E2), and to a minor extent, that of androgen 4-androstenedione to testosterone
Background:	dehydrogenase type 1, is a member of 17-beta-hydroxysteroid dehydrogenases that compose a family of 14 enzymes and catalyze the conversion between the low active 17-ketosteroids and the highly active 17 beta-hydroxysteroids. Human HSD17B1 contains 327 amino acids and exists as a homodimer with two identical subunits of 34.5 kDa (1-2). It uses NADPH as a cofactor to predominantly catalyze the reduction of inactive estrone (E1) to the active estrogen 17 beta-estradiol (E2), and to a minor extent, that of androgen 4-androstenedione to testosterone (3). The enzyme can regulate intracellular concentrations of active sex steroids and plays a
	dehydrogenase type 1, is a member of 17-beta-hydroxysteroid dehydrogenases that compose a family of 14 enzymes and catalyze the conversion between the low active 17-ketosteroids and the highly active 17 beta-hydroxysteroids. Human HSD17B1 contains 327 amino acids and exists as a homodimer with two identical subunits of 34.5 kDa (1-2). It uses NADPH as a cofactor to predominantly catalyze the reduction of inactive estrone (E1) to the active estrogen 17 beta-estradiol (E2), and to a minor extent, that of androgen 4-androstenedione to testosterone (3). The enzyme can regulate intracellular concentrations of active sex steroids and plays a major role in determining the gradient between the E2 concentrations in serum and peripheral
Gene ID:	dehydrogenase type 1, is a member of 17-beta-hydroxysteroid dehydrogenases that compose a family of 14 enzymes and catalyze the conversion between the low active 17-ketosteroids and the highly active 17 beta-hydroxysteroids. Human HSD17B1 contains 327 amino acids and exists as a homodimer with two identical subunits of 34.5 kDa (1-2). It uses NADPH as a cofactor to predominantly catalyze the reduction of inactive estrone (E1) to the active estrogen 17 beta-estradiol (E2), and to a minor extent, that of androgen 4-androstenedione to testosterone (3). The enzyme can regulate intracellular concentrations of active sex steroids and plays a major role in determining the gradient between the E2 concentrations in serum and peripheral tissues (4-6).
Background:  Gene ID:  UniProt:  Pathways:	dehydrogenase type 1, is a member of 17-beta-hydroxysteroid dehydrogenases that compose a family of 14 enzymes and catalyze the conversion between the low active 17-ketosteroids and the highly active 17 beta-hydroxysteroids. Human HSD17B1 contains 327 amino acids and exists as a homodimer with two identical subunits of 34.5 kDa (1-2). It uses NADPH as a cofactor to predominantly catalyze the reduction of inactive estrone (E1) to the active estrogen 17 beta-estradiol (E2), and to a minor extent, that of androgen 4-androstenedione to testosterone (3). The enzyme can regulate intracellular concentrations of active sex steroids and plays a major role in determining the gradient between the E2 concentrations in serum and peripheral tissues (4-6).

Assay Time:	3 h
Plate:	Pre-coated
Protocol:	<ul> <li>Step 1. Add 50 μL of Standard or Sample per well. Incubate 2 hours.</li> <li>Step 2. Wash, then add 50 μL of Biotinylated Antibody per well. Incubate 1 hour.</li> <li>Step 3. Wash, then add 50 μL of SP Conjugate per well. Incubate 30 minutes.</li> <li>Step 4. Wash, then add 50 μL of Chromogen Substrate per well. Incubate 30 minutes.</li> <li>Step 5. Add 50 μL of Stop Solution per well. Read at 450 nm immediately.</li> </ul>
Reagent Preparation:	Freshly dilute all reagents and bring all reagents to room temperature before use. EIA Diluent
	Concentrate (10x): If crystals have formed in the concentrate, mix gently until the crystals have
	completely dissolved. Dilute the EIA Diluent Concentrate 10-fold with reagent grade water to
	produce a 1x solution. Store for up to 30 days at 2-8 °C. 4 Human HSD17B1 Standard:
	Reconstitute the Human HSD17B1 Standard (50 ng) with 0.5 mL of Standard Diluent to
	generate a 100 ng/mL standard stock solution. Allow the vial to sit for 10 minutes with gentle
	agitation prior to making dilutions. Prepare duplicate or triplicate standard points by serially
	diluting from the standard stock solution (100 ng/mL) 2-fold with equal volume of EIA Diluent to
	produce 50, 25, 12.5, 6.25, 3.125, 1.563, and 0.781 ng/mL solutions. EIA Diluent serves as the
	zero standard (0 ng/mL). Aliquot remaining stock solution to limit repeated freeze-thaw cycles.
	This solution should be stored at -20 °C and used within 48 hours. Standard Point Dilution
	[HSD17B1] (ng/mL) P1 1 part Standard (100 ng/mL) + 1 part EIA Diluent 50 P2 1 part P1 + 1
	part EIA Diluent 25 P3 1 part P2 + 1 part EIA Diluent 12.5 P4 1 part P3 + 1 part EIA Diluent 6.25
	P5 1 part P4 + 1 part EIA Diluent 3.125 P6 1 part P5 + 1 part EIA Diluent 1.563 P7 1 part P6 + 1
	part EIA Diluent 0.781 P8 EIA Diluent 0.0 Biotinylated Human HSD17B1 Antibody (50x): Spin
	down the antibody briefly and dilute the desired amount of the antibody 50-fold with EIA Diluent
	to produce a 1x solution. The undiluted antibody should be stored at -20 °C. Wash Buffer
	Concentrate (20x): If crystals have formed in the concentrate, mix gently until the crystals have
	completely dissolved. Dilute the Wash Buffer Concentrate 20-fold with reagent grade water to
	produce a 1x solution. SP Conjugate (100x): Spin down the SP Conjugate briefly and dilute the
	desired amount of the conjugate 100-fold with EIA Diluent to produce a 1x solution. The
	undiluted conjugate should be stored at -20 °C.
Sample Collection:	Plasma: Collect plasma using one-tenth volume of 0.1 M sodium citrate as an anticoagulant.
	Contribute completed 2000 v. a for 10 minutes and collect places. A 2 fold complete dilution is

Plasma: Collect plasma using one-tenth volume of 0.1 M sodium citrate as an anticoagulant. Centrifuge samples at 3000 x g for 10 minutes and collect plasma. A 2-fold sample dilution is suggested into EIA Diluent, however, user should determine optimal dilution factor depending on application needs. The undiluted samples can be stored at -20 °C or below for up to 3 months. Avoid repeated freeze-thaw cycles (EDTA or Heparin can also be used as an anticoagulant). Serum: Samples should be collected into a serum separator tube. After clot

formation, centrifuge samples at 3000 x g for 10 minutes and remove serum. A 2-fold sample dilution is suggested into EIA Diluent, however, user should determine optimal dilution factor depending on application needs. The undiluted samples can be stored at -20  $^{\circ}$ C or below for up to 3 months. Avoid repeated freeze-thaw cycles. Cell Culture Supernatants: Centrifuge cell culture media at 3000 x g for 10 minutes at 4  $^{\circ}$ C to remove debris and collect supernatants. Samples can be stored at -20  $^{\circ}$ C or below. Avoid repeated freeze-thaw cycles.

Assay Procedure:

Prepare all reagents, standard solutions, and samples as instructed. Bring all reagents to room temperature before use. The assay is performed at room temperature (20-25 °C). Remove excess microplate strips from the plate frame and return them immediately to the foil pouch with desiccants inside. Reseal the pouch securely to minimize exposure to water vapor and store in a vacuum desiccator. 5 Add 50 l of Human HSD17B1 Standard or sample to each well. Gently tap plate to thoroughly coat the wells. Break any bubbles that may have formed. Cover wells with a sealing tape and incubate for 2 hours. Start the timer after the last addition. Wash five times with 200 I of Wash Buffer manually. Invert the plate each time and decant the contents, hit 4-5 times on absorbent material to completely remove the liquid. If using a machine, wash six times with 300 l of Wash Buffer and then invert the plate, decanting the contents, hit 4-5 times on absorbent material to completely remove the liquid. Add 50 l of Biotinylated Human HSD17B1 Antibody to each well. Gently tap plate to thoroughly coat the wells. Break any bubbles that may have formed. Cover wells with a sealing tape and incubate for 1 hour. Wash the microplate as described above. Add 50 I of SP Conjugate to each well. Gently tap plate to thoroughly coat the wells. Break any bubbles that may have formed. Cover wells with a sealing tape and incubate for 30 minutes. Turn on the microplate reader and set up the program in advance. Wash the microplate as described above. Add 50 I of Chromogen Substrate to each well. Gently tap plate to thoroughly coat the wells. Break any bubbles that may have formed. Incubate for 30 minutes or until the optimal blue color density develops. Add 50 I of Stop Solution to each well. The color will change from blue to yellow. Gently tap plate to ensure thorough mixing. Break any bubbles that may have formed. Read the absorbance on a microplate reader at a wavelength of 450 nm immediately. If wavelength correction is available, subtract readings at 570 nm from those at 450 nm to correct optical imperfections. Otherwise, read the plate at 450 nm only. Please note that some unstable black particles may be generated at high concentration points after stopping the reaction for about 10 minutes, which will reduce the readings.

Calculation of Results:

- Calculate the mean value of the duplicate or triplicate readings for each standard and sample.
- · To generate a standard curve, plot the graph using the standard concentrations on the x-axis

and the corresponding mean 450 nm absorbance (OD) on the y-axis. The best-fit line can be determined by regression analysis using log-log or four-parameter logistic curve-fit.

• Determine the unknown sample concentration from the standard curve and multiply the value by the dilution factor.

Restrictions:

For Research Use only

#### Handling

Handling Advice:

This product is for Research Use Only and is not intended for use in diagnostic procedures. Prepare all reagents (diluent buffer, wash buffer, standard, biotinylated antibody, and SP conjugate) as instructed, prior to running the assay. Prepare all samples prior to running the assay. The dilution factors for the samples are suggested in this insert. However, the user should determine the optimal dilution factor. 2 Spin down the SP conjugate vial, the biotinylated antibody vial, and the standard diluent vial before opening and using contents. The Stop Solution is an acidic solution. The kit should not be used beyond the expiration date.

Storage:

4 °C,-20 °C

Storage Comment:

Upon arrival, immediately store components of the kit at recommended temperatures up to the expiration date. Store Standard, SP Conjugate, and Biotinylated Antibody at -20°C. Store Microplate, Diluent Concentrate (10x), Standard Diluent (1x), Wash Buffer, Stop Solution, and Chromogen Substrate at 2-8°C. Unused microplate wells may be returned to the foil pouch with the desiccant packs and resealed. May be stored for up to 30 days in a vacuum desiccator.

## **Images**




Image 1.