

## Datasheet for ABIN5651271

### **OASL ELISA Kit**



#### Overview

Quantity:	96 tests
Target:	OASL
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	15.62 pg/mL - 1000 pg/mL
Minimum Detection Limit:	15.62 pg/mL
Application:	ELISA

#### **Product Details**

Sample Type:	Cell Lysate, Plasma, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of 2',5'-Oligoadenylate Synthetase Like Protein (OASL). No significant cross-reactivity or interference between 2',5'-Oligoadenylate Synthetase Like Protein (OASL) and analogues was observed.
Sensitivity:	6.8 pg/mL

## **Target Details**

Target:	OASL
Alternative Name:	2',5'-Oligoadenylate Synthetase Like Protein (OASL Products)

# Target Details Gene Name: 2',5'-Oligoadenylate Synthetase Like Protein Background: Gene Aliases: TRIP14, p590ASL, 59 kDa 2'-5'-oligoadenylate synthase-like protein, Thyroid receptor-interacting protein 14 Gene ID: 8638 UniProt: Q15646 **Application Details** Comment: The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end. Assay Time: 3 h Pre-coated Plate: Protocol: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to 2',5'-Oligoadenylate Synthetase Like Protein (OASL). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to 2',5'-Oligoadenylate Synthetase Like Protein (OASL). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain 2',5'-Oligoadenylate Synthetase Like Protein (OASL), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of 2',5'-Oligoadenylate Synthetase Like Protein (OASL) in the samples is then determined by comparing the O.D. of the samples to the standard curve. Assay Precision: Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level

Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level 2',5'-Oligoadenylate Synthetase Like Protein (OASL) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10%

2',5'-Oligoadenylate Synthetase Like Protein (OASL) were tested 20 times on one plate,

respectively

## **Application Details**

	Inter-Assay: CV<12%
Restrictions:	For Research Use only
Handling	
Handling Advice:	The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.
Storage:	4 °C,-20 °C
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.
Expiry Date:	4-8 months