

Datasheet for ABIN5651278 **BDH1 ELISA Kit**



Overview

Quantity:	96 tests
Target:	BDH1
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.156 ng/mL - 10 ng/mL
Minimum Detection Limit:	0.156 ng/mL
Application:	ELISA

Product Details

Sample Type:	Cell Lysate, Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of 3-Hydroxybutyrate Dehydrogenase 1 (BDH1). No significant cross-reactivity or interference between 3- Hydroxybutyrate Dehydrogenase 1 (BDH1) and analogues was observed.
Sensitivity:	0.055 ng/mL
Target Details	
Target:	BDH1

Alternative Name:	3-Hydroxybutyrate Dehydrogenase 1 (BDH1 Products)

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/3 | Product datasheet for ABIN5651278 | 07/25/2024 | Copyright antibodies-online. All rights reserved.

Target Details	
Background:	Gene Name: 3-Hydroxybutyrate Dehydrogenase 1
	Gene Aliases: DBH, SDR9C1, D-beta-Hydroxybutyrate Dehydrogenase,Mitochondrial, Short
	Chain Dehydrogenase/Reductase Family 9C,Member 1
Gene ID:	622
UniProt:	Q02338
Pathways:	Response to Growth Hormone Stimulus

Application Details

Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.
Assay Time:	3 h
Plate:	Pre-coated
Protocol:	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to 3-Hydroxybutyrate Dehydrogenase 1 (BDH1). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to 3-Hydroxybutyrate Dehydrogenase 1 (BDH1). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain 3-Hydroxybutyrate Dehydrogenase 1 (BDH1), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of 3-Hydroxybutyrate Dehydrogenase 1 (BDH1) in the samples is then determined by comparing the 0.D. of the samples to the standard curve.
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level 3- Hydroxybutyrate Dehydrogenase 1 (BDH1) were tested 20 times on one plate, respectively Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level 3- Hydroxybutyrate Dehydrogenase 1 (BDH1) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10%

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 2/3 | Product datasheet for ABIN5651278 | 07/25/2024 | Copyright antibodies-online. All rights reserved.

Application Details	
	Inter-Assay: CV<12%
Restrictions:	For Research Use only
Handling	
Handling Advice:	The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.
Storage:	4 °C,-20 °C
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.
Expiry Date:	4-8 months