

Datasheet for ABIN5651802 **Angiotensin III ELISA Kit**



[Go to Product page](#)

Overview

Quantity:	96 tests
Target:	Angiotensin III (AGT3)
Reactivity:	Human
Method Type:	Competition ELISA
Detection Range:	6.17 pg/mL - 500 pg/mL
Minimum Detection Limit:	6.17 pg/mL
Application:	ELISA

Product Details

Sample Type:	Plasma, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of Angiotensin III (AngIII). No significant cross-reactivity or interference between Angiotensin III (AngIII) and analogues was observed.
Sensitivity:	2.77 pg/mL

Target Details

Target:	Angiotensin III (AGT3)
Alternative Name:	Angiotensin III (AGT3 Products)

Target Details

Background: Gene Name: Angiotensin III
Gene Aliases: Ang-III, Angiotensin-3

Application Details

Comment: The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

Assay Time: 2 h

Plate: Pre-coated

Protocol: This assay employs the competitive inhibition enzyme immunoassay technique. A monoclonal antibody specific to Angiotensin III (AngIII) has been pre-coated onto a microplate. A competitive inhibition reaction is launched between biotin labeled Angiotensin III (AngIII) and unlabeled Angiotensin III (AngIII) (Standards or samples) with the pre-coated antibody specific to Angiotensin III (AngIII). After incubation the unbound conjugate is washed off. Next, avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. The amount of bound HRP conjugate is reverse proportional to the concentration of Angiotensin III (AngIII) in the sample. After addition of the substrate solution, the intensity of color developed is reverse proportional to the concentration of Angiotensin III (AngIII) in the sample.

Assay Precision: Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Angiotensin III (AngIII) were tested 20 times on one plate, respectively
Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Angiotensin III (AngIII) were tested on 3 different plates, 8 replicates in each plate. $CV(\%) = \frac{SD}{mean} \times 100$
Intra-Assay: $CV < 10\%$
Inter-Assay: $CV < 12\%$

Restrictions: For Research Use only

Handling

Handling Advice: The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all

Handling

steps should be completed without interruption.

Storage: 4 °C,-20 °C

Storage Comment: -20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.

Expiry Date: 4-8 months