

## Datasheet for ABIN5651970 **Aprotinin CLIA Kit**



[Go to Product page](#)

### Overview

Quantity:	96 tests
Target:	Aprotinin
Reactivity:	Cow
Method Type:	Competition ELISA
Detection Range:	1.95 ng/mL - 500 ng/mL
Minimum Detection Limit:	1.95 ng/mL
Application:	ELISA

### Product Details

Sample Type:	Cell Culture Supernatant, Cell Lysate, Plasma, Serum, Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Chemiluminescent
Specificity:	This assay has high sensitivity and excellent specificity for detection of Aprotinin (AP). No significant cross-reactivity or interference between Aprotinin (AP) and analogues was observed.
Sensitivity:	0.71 ng/mL

### Target Details

Target:	Aprotinin
Alternative Name:	Aprotinin ( <a href="#">Aprotinin Products</a> )
Background:	Gene Name: Aprotinin

## Target Details

---

Gene Aliases: BPTI, Trasylol, Pancreatic Trypsin Inhibitor, Basic protease inhibitor

## Application Details

---

Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.
Assay Time:	2 h
Plate:	Pre-coated
Protocol:	The microplate provided in this kit has been pre-coated with a monoclonal antibody specific to Aprotinin (AP). A competitive inhibition reaction is launched between biotin labeled Aprotinin (AP) and unlabeled Aprotinin (AP) (Standards or samples) with the pre-coated antibody specific to Aprotinin (AP). After incubation the unbound conjugate is washed off. Next, avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. The amount of bound HRP conjugate is reverse proportional to the concentration of Aprotinin (AP) in the sample. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is reverse proportional to the Aprotinin (AP) level in the sample or standard.
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Aprotinin (AP) were tested 20 times on one plate, respectively Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Aprotinin (AP) were tested on 3 different plates, 8 replicates in each plate. CV(%) = $\frac{SD}{mean} \times 100$ Intra-Assay: CV<10% Inter-Assay: CV<12%
Restrictions:	For Research Use only

## Handling

---

Handling Advice:	Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.
Storage:	4 °C, -20 °C

## Handling

---

Storage Comment: -20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.

---

Expiry Date: 4-8 months