

Datasheet for ABIN5652089

ATP6AP2 ELISA Kit



_					
	W	0	rv	10	W

Quantity:	96 tests	
Target:	ATP6AP2	
Reactivity:	Human	
Method Type:	Sandwich ELISA	
Detection Range:	0.312 ng/mL - 20 ng/mL	
Minimum Detection Limit:	0.312 ng/mL	
Application:	ELISA	

Product Details

Sample Type:	Cell Lysate, Tissue Homogenate	
Analytical Method:	Quantitative	
Detection Method:	Colorimetric	
Specificity:	This assay has high sensitivity and excellent specificity for detection of ATPase, H+ Transporting, Lysosomal Accessory Protein 2 (ATP6AP2). No significant cross-reactivity or interference between ATPase, H+ Transporting, Lysosomal Accessory Protein 2 (ATP6AP2) and analogues was observed.	
Sensitivity:	0.115 ng/mL	

Target Details

Target:	ATP6AP2
Alternative Name:	ATPase, H+ Transporting, Lysosomal Accessory Protein 2 (ATP6AP2 Products)

Target Details

Background:	Gene Name: ATPase, H+ Transporting, Lysosomal Accessory Protein 2	
	Gene Aliases: ATP6IP2, CAPER, ELDF10, XMRE, Renin/prorenin receptor, Embryonic liver differentiation factor 10, Vacuolar ATP synthase membrane sector-associated protein M8-9	
Cons ID:	<u> </u>	
Gene ID:	10159	
UniProt:	075787	
Pathways:	ACE Inhibitor Pathway, Peptide Hormone Metabolism, Regulation of Systemic Arterial Blood	
	Pressure by Hormones	
Application Details		
Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than	
	5 % within the expiration date under appropriate storage condition. To minimize extra influence	
	on the performance, operation procedures and lab conditions, especially room temperature, air	
	humidity, incubator temperature should be strictly controlled. It is also strongly suggested that	
	the whole assay is performed by the same operator from the beginning to the end.	
Assay Time:	3 h	
Plate:	Pre-coated	
Protocol:	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate	
	provided in this kit has been pre-coated with an antibody specific to ATPase, H+ Transporting,	
	Lysosomal Accessory Protein 2 (ATP6AP2). Standards or samples are then added to the	
	appropriate microtiter plate wells with a biotin-conjugated antibody specific to ATPase, H+	
	Transporting, Lysosomal Accessory Protein 2 (ATP6AP2). Next, Avidin conjugated to	
	Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB	
	substrate solution is added, only those wells that contain ATPase, H+ Transporting, Lysosomal	
	Accessory Protein 2 (ATP6AP2), biotin-conjugated antibody and enzyme-conjugated Avidin will	
	exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of	
	sulphuric acid solution and the color change is measured spectrophotometrically at a	
	wavelength of 450nm \pm 10nm. The concentration of ATPase, H+ Transporting, Lysosomal	
	Accessory Protein 2 (ATP6AP2) in the samples is then determined by comparing the O.D. of the	
	samples to the standard curve.	
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level	
	ATPase, H+ Transporting, Lysosomal Accessory Protein 2 (ATP6AP2) were tested 20 times on	
	one plate, respectively	
	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level	

Application Details

	ATPase, H+ Transporting, Lysosomal Accessory Protein 2 (ATP6AP2) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10% Inter-Assay: CV<12%	
Restrictions:	For Research Use only	
Handling		
Handling Advice:	The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.	
Storage:	4 °C,-20 °C	
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.	
Expiry Date:	4-8 months	