

## Datasheet for ABIN5652097

# **ATP1B1 ELISA Kit**



#### Overview

Quantity:	96 tests
Target:	ATP1B1
Reactivity:	Mouse
Method Type:	Sandwich ELISA
Detection Range:	0.156 ng/mL - 10 ng/mL
Minimum Detection Limit:	0.156 ng/mL
Application:	ELISA

#### **Product Details**

Sample Type:	Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of ATPase, Na+/K+ Transporting Beta 1 Polypeptide (ATP1b1). No significant cross-reactivity or interference between ATPase, Na+/K+ Transporting Beta 1 Polypeptide (ATP1b1) and analogues was observed.
Sensitivity:	0.071 ng/mL

## Target Details

Target:	ATP1B1
Alternative Name:	ATPase, Na+/K+ Transporting Beta 1 Polypeptide (ATP1B1 Products)

### **Target Details**

Background:	Gene Name: ATPase, Na+/K+ Transporting Beta 1 Polypeptide
	Gene Aliases: ATP1B, Sodium/potassium-dependent ATPase subunit beta-1
Gene ID:	11931
UniProt:	Q545P0
Pathways:	Thyroid Hormone Synthesis, Ribonucleoside Biosynthetic Process, SARS-CoV-2 Protein Interactome
Application Details	
Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.
Assay Time:	3 h
Plate:	Pre-coated
Protocol:	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to ATPase, Na+/K+ Transporting Beta 1 Polypeptide (ATP1b1). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to ATPase, Na+/K+ Transporting Beta 1 Polypeptide (ATP1b1). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain ATPase, Na+/K+ Transporting Beta 1 Polypeptide (ATP1b1), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the colo change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of ATPase, Na+/K+ Transporting Beta 1 Polypeptide (ATP1b1) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level ATPase, Na+/K+ Transporting Beta 1 Polypeptide (ATP1b1) were tested 20 times on one plate, respectively  Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level ATPase, Na+/K+ Transporting Beta 1 Polypeptide (ATP1b1) were tested on 3 different plates, 8

replicates in each plate. CV(%) = SD/meanX100

## **Application Details**

	Intra-Assay: CV<10%
	Inter-Assay: CV<12%
Restrictions:	For Research Use only
Handling	
Handling Advice:	The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and
	specimen samples should be assayed in duplicate. Once the procedure has been started, all
	steps should be completed without interruption.
Storage:	4 °C,-20 °C
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at
	4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant
	pack. Minimize freeze/thaw cycles.
Expiry Date:	4-8 months