Datasheet for ABIN5652403

C-Peptide CLIA Kit

Go to Product page

Overview

Quantity:	96 tests
Target:	C-Peptide
Reactivity:	Human
Method Type:	Competition ELISA
Detection Range:	78.12 pg/mL - 20000 pg/mL
Minimum Detection Limit:	78.12 pg/mL
Application:	ELISA

Product Details

Sample Type:	Cell Culture Supernatant, Cell Lysate, Plasma, Serum, Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Chemiluminescent
Specificity:	This assay has high sensitivity and excellent specificity for detection of C-Peptide (CP). No significant cross-reactivity or interference between C-Peptide (CP) and analogues was observed.
Sensitivity:	31.5 pg/mL

Target Details

Target:	C-Peptide
Abstract:	C-Peptide Products

Target Details

Background:	Gene Name: C-Peptide
Gene ID:	3630
UniProt:	P01308

Application Details

Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.
Assay Time:	2 h
Plate:	Pre-coated
Protocol:	The microplate provided in this kit has been pre-coated with a monoclonal antibody specific to C-Peptide (CP). A competitive inhibition reaction is launched between biotin labeled C-Peptide (CP) and unlabeled C-Peptide (CP) (Standards or samples) with the pre-coated antibody specific to C-Peptide (CP). After incubation the unbound conjugate is washed off. Next, avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. The amount of bound HRP conjugate is reverse proportional to the concentration of C-Peptide (CP) in the sample. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is reverse proportional to the C-Peptide (CP) level in the sample or standard.
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level C-Peptide (CP) were tested 20 times on one plate, respectively Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level C-Peptide (CP) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10% Inter-Assay: CV<12%
Restrictions:	For Research Use only

Handling

Handling Advice:	Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed

Handling

	without interruption.
Storage:	4 °C,-20 °C
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.
Expiry Date:	4-8 months

