

Datasheet for ABIN5652439 C1QTNF1 ELISA Kit



Go to Product page

Overview

Quantity:	96 tests
Target:	C1QTNF1
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	1.56 ng/mL - 100 ng/mL
Minimum Detection Limit:	1.56 ng/mL
Application:	ELISA

Product Details

Sample Type:	Plasma, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of C1q And Tumor Necrosis Factor Related Protein 1 (C1QTNF1). No significant cross-reactivity or interference between C1q And Tumor Necrosis Factor Related Protein 1 (C1QTNF1) and analogues was observed.
Sensitivity:	0.65 ng/mL
Target Details	

Target: C1QTNF1 Alternative Name: C1q And Tumor Necrosis Factor Related Protein 1 (C1QTNF1 Products)

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/3 | Product datasheet for ABIN5652439 | 07/25/2024 | Copyright antibodies-online. All rights reserved.

Target Details	
Background:	Gene Name: C1q And Tumor Necrosis Factor Related Protein 1
	Gene Aliases: GIP, CTRP1, ZSIG37, G Protein Coupled Receptor Interacting Protein
Gene ID:	114897
UniProt:	Q9BXJ1
Application Details	
Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than
	5 % within the expiration date under appropriate storage condition. To minimize extra influence
	on the performance, operation procedures and lab conditions, especially room temperature, air
	humidity, incubator temperature should be strictly controlled. It is also strongly suggested that
	the whole assay is performed by the same operator from the beginning to the end.
Assay Time:	3 h
Plate:	Pre-coated
Protocol:	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate
	provided in this kit has been pre-coated with an antibody specific to C1q And Tumor Necrosis
	Factor Related Protein 1 (C1QTNF1). Standards or samples are then added to the appropriate
	microtiter plate wells with a biotin-conjugated antibody specific to C1q And Tumor Necrosis
	Factor Related Protein 1 (C1QTNF1). Next, Avidin conjugated to Horseradish Peroxidase (HRP)
	is added to each microplate well and incubated. After TMB substrate solution is added, only
	those wells that contain C1q And Tumor Necrosis Factor Related Protein 1 (C1QTNF1), biotin-
	conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-
	substrate reaction is terminated by the addition of sulphuric acid solution and the color change
	is measured spectrophotometrically at a wavelength of 450nm \pm 10nm. The concentration of
	C1q And Tumor Necrosis Factor Related Protein 1 (C1QTNF1) in the samples is then
	determined by comparing the O.D. of the samples to the standard curve.
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level C1c
	And Tumor Necrosis Factor Related Protein 1 (C1QTNF1) were tested 20 times on one plate,
	respectively
	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level
	C1q And Tumor Necrosis Factor Related Protein 1 (C1QTNF1) were tested on 3 different plates
	8 replicates in each plate. CV(%) = SD/meanX100
	Intra-Assay: CV<10%
	Inter-Assay: CV<12%

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 2/3 | Product datasheet for ABIN5652439 | 07/25/2024 | Copyright antibodies-online. All rights reserved.

Application Details	
Restrictions:	For Research Use only
Handling	
Handling Advice:	The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.
Storage:	4 °C,-20 °C
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.
Expiry Date:	4-8 months