

Datasheet for ABIN5652467

CALB1 ELISA Kit



Overview

Quantity:	96 tests
Target:	CALB1
Reactivity:	Rat
Method Type:	Sandwich ELISA
Detection Range:	9.38 pg/mL - 600 pg/mL
Minimum Detection Limit:	9.38 pg/mL
Application:	ELISA

Product Details

Sample Type:	Plasma, Serum, Tissue Homogenate, Urine
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of Calbindin (CALB). No significant cross-reactivity or interference between Calbindin (CALB) and analogues was observed.
Sensitivity:	3.3 pg/mL

Target Details

Target:	CALB1
Alternative Name:	Calbindin (CALB1 Products)

Target Details

Background:	Gene Name: Calbindin
	Gene Aliases: CAB27, Calbindin 1,28 kDa, Calbindin D28, Vitamin D-dependent calcium-binding
	protein, avian-type
Gene ID:	83839
UniProt:	P07171
Application Details	
Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than
	5 % within the expiration date under appropriate storage condition. To minimize extra influence
	on the performance, operation procedures and lab conditions, especially room temperature, air
	humidity, incubator temperature should be strictly controlled. It is also strongly suggested that
	the whole assay is performed by the same operator from the beginning to the end.
Assay Time:	3 h
Plate:	Pre-coated
Protocol:	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate
	provided in this kit has been pre-coated with an antibody specific to Calbindin (CALB).
	Standards or samples are then added to the appropriate microtiter plate wells with a biotin-
	conjugated antibody specific to Calbindin (CALB). Next, Avidin conjugated to Horseradish
	Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution
	is added, only those wells that contain Calbindin (CALB), biotin-conjugated antibody and
	enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is
	terminated by the addition of sulphuric acid solution and the color change is measured
	spectrophotometrically at a wavelength of 450nm \pm 10nm. The concentration of Calbindin
	(CALB) in the samples is then determined by comparing the O.D. of the samples to the standard
	curve.
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level
	Calbindin (CALB) were tested 20 times on one plate, respectively
	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level
	Calbindin (CALB) were tested on 3 different plates, 8 replicates in each plate. CV(%) =
	SD/meanX100
	Intra-Assay: CV<10%
	Inter-Assay: CV<12%
Restrictions:	For Research Use only

Handling

Handling Advice:	The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.
Storage:	4 °C,-20 °C
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.
Expiry Date:	4-8 months