

## Datasheet for ABIN5652487 **CACNA1S ELISA Kit**



[Go to Product page](#)

### Overview

Quantity:	96 tests
Target:	CACNA1S
Reactivity:	Mouse
Method Type:	Sandwich ELISA
Detection Range:	0.312 ng/mL - 20 ng/mL
Minimum Detection Limit:	0.312 ng/mL
Application:	ELISA

### Product Details

Sample Type:	Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of Calcium Channel, Voltage Dependent, L-Type, Alpha 1S Subunit (CACNa1S). No significant cross-reactivity or interference between Calcium Channel, Voltage Dependent, L-Type, Alpha 1S Subunit (CACNa1S) and analogues was observed.
Sensitivity:	0.101 ng/mL

### Target Details

Target:	CACNA1S
Alternative Name:	Calcium Channel, Voltage Dependent, L-Type, Alpha 1S Subunit ( <a href="#">CACNA1S Products</a> )

## Target Details

---

Background: Gene Name: Calcium Channel, Voltage Dependent, L-Type, Alpha 1S Subunit  
Gene Aliases: CACNL1A3, CCHL1A3, Cav1.1, HOKPP, MHS5, hypoPP, DHPR, Dihydropyridine Receptor, Voltage-gated calcium channel subunit alpha Cav1.1

UniProt: [Q02789](#)

Pathways: [Skeletal Muscle Fiber Development](#)

## Application Details

---

Comment: The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

Assay Time: 3 h

Plate: Pre-coated

Protocol: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Calcium Channel, Voltage Dependent, L-Type, Alpha 1S Subunit (CACNa1S). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Calcium Channel, Voltage Dependent, L-Type, Alpha 1S Subunit (CACNa1S). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Calcium Channel, Voltage Dependent, L-Type, Alpha 1S Subunit (CACNa1S), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Calcium Channel, Voltage Dependent, L-Type, Alpha 1S Subunit (CACNa1S) in the samples is then determined by comparing the O.D. of the samples to the standard curve.

Assay Precision: Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Calcium Channel, Voltage Dependent, L-Type, Alpha 1S Subunit (CACNa1S) were tested 20 times on one plate, respectively  
Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Calcium Channel, Voltage Dependent, L-Type, Alpha 1S Subunit (CACNa1S) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100

## Application Details

---

Intra-Assay: CV<10%

Inter-Assay: CV<12%

---

Restrictions: For Research Use only

---

## Handling

---

Handling Advice: The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.

---

Storage: 4 °C,-20 °C

---

Storage Comment: -20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.

---

Expiry Date: 4-8 months

---