

## Datasheet for ABIN5652742 **SIVA1 ELISA Kit**



[Go to Product page](#)

### Overview

Quantity:	96 tests
Target:	SIVA1
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.156 ng/mL - 10 ng/mL
Minimum Detection Limit:	0.156 ng/mL
Application:	ELISA

### Product Details

Sample Type:	Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of CD27 Binding Protein (CD27BP). No significant cross-reactivity or interference between CD27 Binding Protein (CD27BP) and analogues was observed.
Sensitivity:	0.068 ng/mL

### Target Details

Target:	SIVA1
Alternative Name:	CD27 Binding Protein ( <a href="#">SIVA1 Products</a> )

## Target Details

---

Background:	Gene Name: CD27 Binding Protein Gene Aliases: SIVA, SIVA1, Apoptosis-Inducing Factor, Apoptosis regulatory protein Siva
Gene ID:	10572
UniProt:	<a href="#">O15304</a>
Pathways:	<a href="#">Positive Regulation of Endopeptidase Activity</a>

## Application Details

---

Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.
Assay Time:	3 h
Plate:	Pre-coated
Protocol:	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to CD27 Binding Protein (CD27BP). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to CD27 Binding Protein (CD27BP). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain CD27 Binding Protein (CD27BP), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of $450\text{nm} \pm 10\text{nm}$ . The concentration of CD27 Binding Protein (CD27BP) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level CD27 Binding Protein (CD27BP) were tested 20 times on one plate, respectively Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level CD27 Binding Protein (CD27BP) were tested on 3 different plates, 8 replicates in each plate. $CV(\%) = SD/\text{mean} \times 100$ Intra-Assay: $CV < 10\%$ Inter-Assay: $CV < 12\%$
Restrictions:	For Research Use only

## Handling

---

Handling Advice:	The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.
Storage:	4 °C,-20 °C
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.
Expiry Date:	4-8 months