

## Datasheet for ABIN5652954 **CTRB1 ELISA Kit**



[Go to Product page](#)

### Overview

Quantity:	96 tests
Target:	CTRB1
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	1.56 ng/mL - 100 ng/mL
Minimum Detection Limit:	1.56 ng/mL
Application:	ELISA

### Product Details

Sample Type:	Plasma, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of Chymotrypsinogen B1 (CTRB1). No significant cross-reactivity or interference between Chymotrypsinogen B1 (CTRB1) and analogues was observed.
Sensitivity:	0.68 ng/mL

### Target Details

Target:	CTRB1
Alternative Name:	Chymotrypsinogen B1 ( <a href="#">CTRB1 Products</a> )

## Target Details

---

Background: Gene Name: Chymotrypsinogen B1

Gene ID: 1504

UniProt: [P17538](#)

## Application Details

---

**Comment:** The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

**Assay Time:** 3 h

**Plate:** Pre-coated

**Protocol:** The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Chymotrypsinogen B1 (CTRB1). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Chymotrypsinogen B1 (CTRB1). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Chymotrypsinogen B1 (CTRB1), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of  $450\text{nm} \pm 10\text{nm}$ . The concentration of Chymotrypsinogen B1 (CTRB1) in the samples is then determined by comparing the O.D. of the samples to the standard curve.

**Assay Precision:** Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Chymotrypsinogen B1 (CTRB1) were tested 20 times on one plate, respectively  
Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Chymotrypsinogen B1 (CTRB1) were tested on 3 different plates, 8 replicates in each plate.  
 $CV(\%) = SD/\text{mean} \times 100$   
Intra-Assay:  $CV < 10\%$   
Inter-Assay:  $CV < 12\%$

**Restrictions:** For Research Use only

## Handling

---

Handling Advice:	The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.
Storage:	4 °C,-20 °C
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.
Expiry Date:	4-8 months