

Datasheet for ABIN5653213 **C1S CLIA Kit**



[Go to Product page](#)

Overview

Quantity: 96 tests

Target: C1S

Reactivity: Human

Method Type: Sandwich ELISA

Detection Range: 468.75 pg/mL - 30000 pg/mL

Minimum Detection Limit: 468.75 pg/mL

Application: ELISA

Product Details

Sample Type: Plasma, Serum

Analytical Method: Quantitative

Detection Method: Chemiluminescent

Specificity: This assay has high sensitivity and excellent specificity for detection of Complement Component 1, S Subcomponent (C1s). No significant cross-reactivity or interference between Complement Component 1, S Subcomponent (C1s) and analogues was observed.

Sensitivity: 18.22 pg/mL

Target Details

Target: C1S

Alternative Name: Complement Component 1, S Subcomponent ([C1S Products](#))

Target Details

Background:	Gene Name: Complement Component 1, S Subcomponent Gene Aliases: C1 esterase
Gene ID:	716
UniProt:	P09871
Pathways:	Complement System

Application Details

Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.
Assay Time:	2 - 3 h
Plate:	Pre-coated
Protocol:	The microplate provided in this kit has been pre-coated with an antibody specific to Complement Component 1, S Subcomponent (C1s). Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to Complement Component 1, S Subcomponent (C1s). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is proportional to the Complement Component 1, S Subcomponent (C1s) level in the sample or standard.,
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Complement Component 1, S Subcomponent (C1s) were tested 20 times on one plate, respectively Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Complement Component 1, S Subcomponent (C1s) were tested on 3 different plates, 8 replicates in each plate. $CV(\%) = SD/mean \times 100$ Intra-Assay: $CV < 10\%$ Inter-Assay: $CV < 12\%$
Restrictions:	For Research Use only

Handling

Handling Advice:	Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.
Storage:	4 °C,-20 °C
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.
Expiry Date:	4-8 months