

# Datasheet for ABIN5653398

# **CDCP1 ELISA Kit**



_					
	W	0	rv	10	W

Quantity:	96 tests
Target:	CDCP1
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.312 ng/mL - 20 ng/mL
Minimum Detection Limit:	0.312 ng/mL
Application:	ELISA

### **Product Details**

Sample Type:	Cell Lysate, Tissue Homogenate	
Analytical Method:	Quantitative	
Detection Method:	Colorimetric	
Specificity:	This assay has high sensitivity and excellent specificity for detection of CUB Domain Containing Protein 1 (CDCP1). No significant cross-reactivity or interference between CUB Domain Containing Protein 1 (CDCP1) and analogues was observed.	
Sensitivity:	0.124 ng/mL	

# Target Details

Target:	CDCP1	
Alternative Name:	CUB Domain Containing Protein 1 (CDCP1 Products)	

# Target Details

Background:	Gene Name: CUB Domain Containing Protein 1	
backyrouriu.	Gene Aliases: CD318, CDCP1, SIMA135, TRASK, Membrane glycoprotein gp140,	
	Transmembrane and associated with src kinases, Subtractive immunization M plus HEp3-	
	associated 135 kDa	
Gene ID:	64866	
UniProt:	Q9H5V8	
Application Details		
Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than	
	5% within the expiration date under appropriate storage condition. To minimize extra influence	
	on the performance, operation procedures and lab conditions, especially room temperature, air	
	humidity, incubator temperature should be strictly controlled. It is also strongly suggested that	
	the whole assay is performed by the same operator from the beginning to the end.	
Assay Time:	3 h	
Plate:	Pre-coated	
Protocol:	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate	
	provided in this kit has been pre-coated with an antibody specific to CUB Domain Containing	
	Protein 1 (CDCP1). Standards or samples are then added to the appropriate microtiter plate	
	wells with a biotin-conjugated antibody specific to CUB Domain Containing Protein 1 (CDCP1).	
	Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and	
	incubated. After TMB substrate solution is added, only those wells that contain CUB Domain	
	Containing Protein 1 (CDCP1), biotin-conjugated antibody and enzyme-conjugated Avidin will	
	exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of	
	sulphuric acid solution and the color change is measured spectrophotometrically at a	
	wavelength of 450nm ± 10nm. The concentration of CUB Domain Containing Protein 1 (CDCP1)	
	in the samples is then determined by comparing the O.D. of the samples to the standard curve.	
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level	
	CUB Domain Containing Protein 1 (CDCP1) were tested 20 times on one plate, respectively	
	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level	
	CUB Domain Containing Protein 1 (CDCP1) were tested on 3 different plates, 8 replicates in	
	each plate. CV(%) = SD/meanX100	
	Intra-Assay: CV<10%	
	Inter-Assay: CV<12%	

# **Application Details**

Restrictions:	For Research Use only	
Handling		
Handling Advice:	The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and	
	specimen samples should be assayed in duplicate. Once the procedure has been started, all	
	steps should be completed without interruption.	
Storage:	4 °C,-20 °C	
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at	
	4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant	
	pack. Minimize freeze/thaw cycles.	
Expiry Date:	4-8 months	