

Datasheet for ABIN5653551

DDT ELISA Kit



Go to Product page

| Overviev | |
|----------|--|
| | |

| Quantity: | 96 tests |
|--------------------------|------------------------|
| Target: | DDT |
| Reactivity: | Human |
| Method Type: | Sandwich ELISA |
| Detection Range: | 0.312 ng/mL - 20 ng/mL |
| Minimum Detection Limit: | 0.312 ng/mL |
| Application: | ELISA |

Product Details

| Sample Type: | Cell Lysate, Tissue Homogenate |
|--------------------|---|
| Analytical Method: | Quantitative |
| Detection Method: | Colorimetric |
| Specificity: | This assay has high sensitivity and excellent specificity for detection of D-Dopachrome Tautomerase (DDT). No significant cross-reactivity or interference between D-Dopachrome Tautomerase (DDT) and analogues was observed. |
| Sensitivity: | 0.117 ng/mL |
| Target Details | |

Target Details

| Target: | DDT |
|-------------------|---|
| Alternative Name: | D-Dopachrome Tautomerase (DDT Products) |

Target Details

| Background: | Gene Name: D-Dopachrome Tautomerase |
|---------------------|---|
| | Gene Aliases: DDCT, D-Dopachrome Decarboxylase, Phenylpyruvate tautomerase II |
| Gene ID: | 1652 |
| UniProt: | P30046 |
| Application Details | |
| Comment: | The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than |
| | 5 % within the expiration date under appropriate storage condition. To minimize extra influence |
| | on the performance, operation procedures and lab conditions, especially room temperature, air |
| | humidity, incubator temperature should be strictly controlled. It is also strongly suggested that |
| | the whole assay is performed by the same operator from the beginning to the end. |
| Assay Time: | 3 h |
| Plate: | Pre-coated |
| Protocol: | The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate |
| | provided in this kit has been pre-coated with an antibody specific to D-Dopachrome |
| | Tautomerase (DDT). Standards or samples are then added to the appropriate microtiter plate |
| | wells with a biotin-conjugated antibody specific to D-Dopachrome Tautomerase (DDT). Next, |
| | Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and |
| | incubated. After TMB substrate solution is added, only those wells that contain D-Dopachrome |
| | Tautomerase (DDT), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a |
| | change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid |
| | solution and the color change is measured spectrophotometrically at a wavelength of 450nm |
| | 10nm. The concentration of D-Dopachrome Tautomerase (DDT) in the samples is then |
| | determined by comparing the O.D. of the samples to the standard curve. |
| Assay Precision: | Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level D- |
| | Dopachrome Tautomerase (DDT) were tested 20 times on one plate, respectively |
| | Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level D- |
| | Dopachrome Tautomerase (DDT) were tested on 3 different plates, 8 replicates in each plate. |
| | CV(%) = SD/meanX100 |
| | Intra-Assay: CV<10% |
| | Inter-Assay: CV<12% |
| Restrictions: | For Research Use only |
| | |

Handling

| Handling Advice: | The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption. |
|------------------|--|
| Storage: | 4 °C,-20 °C |
| Storage Comment: | -20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles. |
| Expiry Date: | 4-8 months |