

## Datasheet for ABIN5653574 **DEFA6 ELISA Kit**



Overview

Quantity:	96 tests
Target:	DEFA6
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.469 ng/mL - 30 ng/mL
Minimum Detection Limit:	0.469 ng/mL
Application:	ELISA

## Product Details

Sample Type:	Cell Lysate, Plasma, Serum, Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of Defensin Alpha 6, Paneth Cell Specific (DEFa6). No significant cross-reactivity or interference between Defensin Alpha 6, Paneth Cell Specific (DEFa6) and analogues was observed.
Sensitivity:	0.21 ng/mL
Target Details	
Target:	DEFA6

Alternative Name:	Defensin Alpha 6, Paneth Cell Specific (DEFA6 Products)

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/3 | Product datasheet for ABIN5653574 | 07/25/2024 | Copyright antibodies-online. All rights reserved.

Background:	Gene Name: Defensin Alpha 6, Paneth Cell Specific
	Gene Aliases: DEF6, Defensin-6
Gene ID:	1671
UniProt:	Q01524
Pathways:	Cellular Response to Molecule of Bacterial Origin
Application Details	
Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than
	5 % within the expiration date under appropriate storage condition. To minimize extra influence
	on the performance, operation procedures and lab conditions, especially room temperature, air
	humidity, incubator temperature should be strictly controlled. It is also strongly suggested that
	the whole assay is performed by the same operator from the beginning to the end.
Assay Time:	3 h
Plate:	Pre-coated
Protocol:	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate
	provided in this kit has been pre-coated with an antibody specific to Defensin Alpha 6, Paneth
	Cell Specific (DEFa6). Standards or samples are then added to the appropriate microtiter plate
	wells with a biotin-conjugated antibody specific to Defensin Alpha 6, Paneth Cell Specific
	(DEFa6). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate
	well and incubated. After TMB substrate solution is added, only those wells that contain
	Defensin Alpha 6, Paneth Cell Specific (DEFa6), biotin-conjugated antibody and enzyme-
	conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by
	the addition of sulphuric acid solution and the color change is measured
	spectrophotometrically at a wavelength of 450nm $\pm$ 10nm. The concentration of Defensin
	Alpha 6, Paneth Cell Specific (DEFa6) in the samples is then determined by comparing the O.D.
	of the samples to the standard curve.
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level
	Defensin Alpha 6, Paneth Cell Specific (DEFa6) were tested 20 times on one plate, respectively
	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level
	Defensin Alpha 6, Paneth Cell Specific (DEFa6) were tested on 3 different plates, 8 replicates in
	each plate. CV(%) = SD/meanX100
	Intra-Assay: CV<10%
	Inter-Assay: CV<12%

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 2/3 | Product datasheet for ABIN5653574 | 07/25/2024 | Copyright antibodies-online. All rights reserved.

Application Details	
Restrictions:	For Research Use only
Handling	
Handling Advice:	The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.
Storage:	4 °C,-20 °C
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.
Expiry Date:	4-8 months