

Datasheet for ABIN5653576 **beta Defensin 1 ELISA Kit**



Overview

| Quantity: | 96 tests |
|--------------------------|-------------------------|
| Target: | beta Defensin 1 (DEFB1) |
| Reactivity: | Cow |
| Method Type: | Sandwich ELISA |
| Detection Range: | 0.156 ng/mL - 10 ng/mL |
| Minimum Detection Limit: | 0.156 ng/mL |
| Application: | ELISA |

Product Details

| Sample Type: | Plasma, Serum, Tissue Homogenate |
|--------------------|---|
| Analytical Method: | Quantitative |
| Detection Method: | Colorimetric |
| Specificity: | This assay has high sensitivity and excellent specificity for detection of Defensin Beta 1 (DEFb1). No significant cross-reactivity or interference between Defensin Beta 1 (DEFb1) and analogues was observed. |
| Sensitivity: | 0.055 ng/mL |
| Target Details | |
| Target: | beta Defensin 1 (DEFB1) |
| Alternative Name: | Defensin Beta 1 (DEFB1 Products) |

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/3 | Product datasheet for ABIN5653576 | 07/25/2024 | Copyright antibodies-online. All rights reserved.

Background:Gene Name: Defensin Beta 1Gene Aliases: B-DF1, DEFB1, BD1, DEF-B1, DEFB101, HBD1

Application Details

| Comment: | The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than |
|------------------|---|
| | 5 % within the expiration date under appropriate storage condition. To minimize extra influence |
| | on the performance, operation procedures and lab conditions, especially room temperature, air |
| | humidity, incubator temperature should be strictly controlled. It is also strongly suggested that |
| | the whole assay is performed by the same operator from the beginning to the end. |
| Assay Time: | 3 h |
| Plate: | Pre-coated |
| Protocol: | The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate |
| | provided in this kit has been pre-coated with an antibody specific to Defensin Beta 1 (DEFb1). |
| | Standards or samples are then added to the appropriate microtiter plate wells with a biotin- |
| | conjugated antibody specific to Defensin Beta 1 (DEFb1). Next, Avidin conjugated to |
| | Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB |
| | substrate solution is added, only those wells that contain Defensin Beta 1 (DEFb1), biotin- |
| | conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme- |
| | substrate reaction is terminated by the addition of sulphuric acid solution and the color change |
| | is measured spectrophotometrically at a wavelength of 450nm \pm 10nm. The concentration of |
| | Defensin Beta 1 (DEFb1) in the samples is then determined by comparing the O.D. of the |
| | samples to the standard curve. |
| Assay Precision: | Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level |
| | Defensin Beta 1 (DEFb1) were tested 20 times on one plate, respectively |
| | Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level |
| | Defensin Beta 1 (DEFb1) were tested on 3 different plates, 8 replicates in each plate. $CV(\%)$ = |
| | SD/meanX100 |
| | Intra-Assay: CV<10% |
| | Inter-Assay: CV<12% |
| Restrictions: | For Research Use only |
| Handling | |
| Handling Advice: | The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and |
| | |

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 2/3 | Product datasheet for ABIN5653576 | 07/25/2024 | Copyright antibodies-online. All rights reserved.

| | specimen samples should be assayed in duplicate. Once the procedure has been started, all |
|------------------|--|
| | steps should be completed without interruption. |
| Storage: | 4 °C,-20 °C |
| Storage Comment: | -20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles. |
| Expiry Date: | 4-8 months |