

Datasheet for ABIN5653716 **DDOST ELISA Kit**



Overview

Quantity:	96 tests
Target:	DDOST
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.312 ng/mL - 20 ng/mL
Minimum Detection Limit:	0.312 ng/mL
Application:	ELISA

Product Details

Sample Type:	Cell Lysate, Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of Dolichyl Diphosphooligosaccharide Protein Glycosyltransferase (DDOST). No significant cross-reactivity or interference between Dolichyl Diphosphooligosaccharide Protein Glycosyltransferase
	(DDOST) and analogues was observed.
Sensitivity:	0.126 ng/mL

Target Details

Target:	DDOST
Alternative Name:	Dolichyl Diphosphooligosaccharide Protein Glycosyltransferase (DDOST Products)

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/3 | Product datasheet for ABIN5653716 | 07/25/2024 | Copyright antibodies-online. All rights reserved.

Target Details	
Background:	Gene Name: Dolichyl Diphosphooligosaccharide Protein Glycosyltransferase Gene Aliases: AGE-R1, OK/SW-cl.45, OST, OST48, WBP1, Dolichyl-diphosphooligosaccharide protein glycosyltransferase 48 kDa subunit, Oligosaccharyl transferase 48 kDa
Gene ID:	1650
UniProt:	P39656
Pathways:	Cell RedoxHomeostasis

Application Details

Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.
Assay Time:	3 h
Plate:	Pre-coated
Protocol:	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Dolichyl Diphosphooligosaccharide Protein Glycosyltransferase (DDOST). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Dolichyl Diphosphooligosaccharide Protein Glycosyltransferase (DDOST). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Dolichyl Diphosphooligosaccharide Protein Glycosyltransferase (DDOST), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Dolichyl Diphosphooligosaccharide Protein Glycosyltransferase (DDOST) in the samples is then determined by comparing the 0.D. of the samples to the standard curve.
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Dolichyl Diphosphooligosaccharide Protein Glycosyltransferase (DDOST) were tested 20 times on one plate, respectively Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Dolichyl Diphosphooligosaccharide Protein Glycosyltransferase (DDOST) were tested on 3

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 2/3 | Product datasheet for ABIN5653716 | 07/25/2024 | Copyright antibodies-online. All rights reserved.

Application Details	
	different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10% Inter-Assay: CV<12%
Restrictions:	For Research Use only
Handling	
Handling Advice:	The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.
Storage:	4 °C,-20 °C
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.
Expiry Date:	4-8 months