

## Datasheet for ABIN5653875

### **EZH2 CLIA Kit**



#### Overview

| Quantity:                | 96 tests                   |
|--------------------------|----------------------------|
| Target:                  | EZH2                       |
| Reactivity:              | Human                      |
| Method Type:             | Sandwich ELISA             |
| Detection Range:         | 156.25 pg/mL - 10000 pg/mL |
| Minimum Detection Limit: | 156.25 pg/mL               |
| Application:             | ELISA                      |

#### **Product Details**

| Sample Type:       | Tissue Homogenate   |
|--------------------|---|
| Analytical Method: | Quantitative  |
| Detection Method:  | Chemiluminescent  |
| Specificity:       | This assay has high sensitivity and excellent specificity for detection of Enhancer Of Zeste Homolog 2 (EZH2). No significant cross-reactivity or interference between Enhancer Of Zeste Homolog 2 (EZH2) and analogues was observed. |
| Sensitivity:       | 5.69 pg/mL  |

## **Target Details**

| Target:           | EZH2  |
|-------------------|---|
| Alternative Name: | Enhancer Of Zeste Homolog 2 (EZH2 Products) |

## Target Details

| Background:         | Gene Name: Enhancer Of Zeste Homolog 2  |
|---------------------|---|
|                     | Gene Aliases: ENX-1, KMT6, KMT6A, Histone-lysine N-methyltransferase EZH2, Lysine N-                    |
|                     | methyltransferase 6   |
| Gene ID:            | 2146  |
| UniProt:            | Q15910  |
| Pathways:           | Retinoic Acid Receptor Signaling Pathway, Regulation of Muscle Cell Differentiation                     |
| Application Details |   |
| Comment:            | The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than |
|                     | 5 % within the expiration date under appropriate storage condition. To minimize extra influence         |
|                     | on the performance, operation procedures and lab conditions, especially room temperature, air           |
|                     | humidity, incubator temperature should be strictly controlled. It is also strongly suggested that       |
|                     | the whole assay is performed by the same operator from the beginning to the end.                        |
| Assay Time:         | 2 - 3 h   |
| Plate:              | Pre-coated  |
| Protocol:           | The microplate provided in this kit has been pre-coated with an antibody specific to Enhancer           |
|                     | Of Zeste Homolog 2 (EZH2). Standards or samples are then added to the appropriate                       |
|                     | microplate wells with a biotin-conjugated antibody specific to Enhancer Of Zeste Homolog 2              |
|                     | (EZH2). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate             |
|                     | well and incubated. Then the mixture of substrate A and B is added to generate glow light               |
|                     | emission kinetics. Upon plate development, the intensity of the emitted light is proportional to        |
|                     | the Enhancer Of Zeste Homolog 2 (EZH2) level in the sample or standard.,                                |
| Assay Precision:    | Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level            |
|                     | Enhancer Of Zeste Homolog 2 (EZH2) were tested 20 times on one plate, respectively                      |
|                     | Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level             |
|                     | Enhancer Of Zeste Homolog 2 (EZH2) were tested on 3 different plates, 8 replicates in each              |
|                     | plate. CV(%) = SD/meanX100  |
|                     | Intra-Assay: CV<10%   |
|                     | Inter-Assay: CV<12%   |
|                     |   |

# Handling

| Handling Advice: | Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be   |
|------------------|--|
|                  | assayed in duplicate. Once the procedure has been started, all steps should be completed   |
|                  | without interruption.  |
| Storage:         | 4 °C,-20 °C  |
| Storage Comment: | -20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles. |
| Expiry Date:     | 4-8 months   |