

Datasheet for ABIN5653996
Estrone CLIA Kit



[Go to Product page](#)

Overview

Quantity:	96 tests
Target:	Estrone
Reactivity:	Various Species
Method Type:	Competition ELISA
Detection Range:	23.44 pg/mL - 6000 pg/mL
Minimum Detection Limit:	23.44 pg/mL
Application:	ELISA

Product Details

Sample Type:	Plasma, Serum
Analytical Method:	Quantitative
Detection Method:	Chemiluminescent
Specificity:	This assay has high sensitivity and excellent specificity for detection of Estrone (E1). No significant cross-reactivity or interference between Estrone (E1) and analogues was observed.
Sensitivity:	7.96 pg/mL

Target Details

Target:	Estrone
Abstract:	Estrone Products
Target Type:	Amino Acid

Target Details

Background: Gene Name: Estrone
Gene Aliases: Oestrone

Application Details

Comment: The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

Assay Time: 2 h

Plate: Pre-coated

Protocol: The microplate provided in this kit has been pre-coated with a monoclonal antibody specific to Estrone (E1). A competitive inhibition reaction is launched between biotin labeled Estrone (E1) and unlabeled Estrone (E1) (Standards or samples) with the pre-coated antibody specific to Estrone (E1). After incubation the unbound conjugate is washed off. Next, avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. The amount of bound HRP conjugate is reverse proportional to the concentration of Estrone (E1) in the sample. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is reverse proportional to the Estrone (E1) level in the sample or standard.

Assay Precision: Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Estrone (E1) were tested 20 times on one plate, respectively
Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Estrone (E1) were tested on 3 different plates, 8 replicates in each plate. $CV(\%) = SD/mean \times 100$
Intra-Assay: $CV < 10\%$
Inter-Assay: $CV < 12\%$

Restrictions: For Research Use only

Handling

Handling Advice: Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.

Storage: 4 °C, -20 °C

Handling

Storage Comment: -20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.

Expiry Date: 4-8 months