# antibodies -online.com





## Datasheet for ABIN5654056

### **FNTA ELISA Kit**



#### Overview

Quantity:	96 tests
Target:	FNTA
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	1.56 ng/mL - 100 ng/mL
Minimum Detection Limit:	1.56 ng/mL
Application:	ELISA

#### **Product Details**

Sample Type:	Plasma, Serum, Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of Farnesyltransferase Alpha (FNTa). No significant cross-reactivity or interference between Farnesyltransferase Alpha (FNTa) and analogues was observed.
Sensitivity:	0.56 ng/mL

## Target Details

Target:	FNTA
Alternative Name:	Farnesyltransferase Alpha (FNTA Products)

## Target Details

Background:	Gene Name: Farnesyltransferase Alpha
	Gene Aliases: FPTA, PTAR2, PGGT1A, Protein Prenyltransferase Alpha Subunit Repeat
	Containing 2, CAAX farnesyltransferase subunit alpha, Ras proteins prenyltransferase subunit
	alpha
Gene ID:	2339
UniProt:	P49354
Pathways:	Response to Water Deprivation, Regulation of G-Protein Coupled Receptor Protein Signaling
Application Details	
Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than
	5 % within the expiration date under appropriate storage condition. To minimize extra influence
	on the performance, operation procedures and lab conditions, especially room temperature, air
	humidity, incubator temperature should be strictly controlled. It is also strongly suggested that
	the whole assay is performed by the same operator from the beginning to the end.
Assay Time:	3 h
Plate:	Pre-coated
Protocol:	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate
	provided in this kit has been pre-coated with an antibody specific to Farnesyltransferase Alpha
	(FNTa). Standards or samples are then added to the appropriate microtiter plate wells with a
	biotin-conjugated antibody specific to Farnesyltransferase Alpha (FNTa). Next, Avidin
	conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated.
	After TMB substrate solution is added, only those wells that contain Farnesyltransferase Alpha
	(FNTa), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color.
	The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the
	color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The
	concentration of Farnesyltransferase Alpha (FNTa) in the samples is then determined by
	comparing the O.D. of the samples to the standard curve.
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level
	Farnesyltransferase Alpha (FNTa) were tested 20 times on one plate, respectively
	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level
	Farnesyltransferase Alpha (FNTa) were tested on 3 different plates, 8 replicates in each plate.
	CV(%) = SD/meanX100
	Intra-Assay: CV<10%

## **Application Details**

	Inter-Assay: CV<12%
Restrictions:	For Research Use only
Handling	
Handling Advice:	The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.
Storage:	4 °C,-20 °C
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.
Expiry Date:	4-8 months