

Datasheet for ABIN5654058 FADD ELISA Kit



Overview

Quantity:	96 tests
Target:	FADD
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.156 ng/mL - 10 ng/mL
Minimum Detection Limit:	0.156 ng/mL
Application:	ELISA

Product Details

Sample Type:	Cell Culture Supernatant, Cell Lysate, Tissue Homogenate	
Analytical Method:	Quantitative	
Detection Method:	Colorimetric	
Specificity:	This assay has high sensitivity and excellent specificity for detection of Fas Associating Death	
	Domain Containing Protein (FADD). No significant cross-reactivity or interference between Fas	
	Associating Death Domain Containing Protein (FADD) and analogues was observed.	

Target Details

Target:	FADD	
Alternative Name:	Fas Associating Death Domain Containing Protein (FADD Products)	
Gene ID:	8772	

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/3 | Product datasheet for ABIN5654058 | 07/25/2024 | Copyright antibodies-online. All rights reserved.

Target Details	
UniProt:	Q13158
Pathways:	Apoptosis, TLR Signaling, Activation of Innate immune Response, Positive Regulation of Endopeptidase Activity, Toll-Like Receptors Cascades
Application Details	
Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.
Assay Time:	3 h
Plate:	Pre-coated
Protocol:	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Fas Associating Death Domain Containing Protein (FADD). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Fas Associating Death Domain Containing Protein (FADD). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Fas Associating Death Domain Containing Protein (FADD) beach antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Fas Associating Death Domain Containing Protein (FADD) in the samples is then determined by comparing the 0.D. of the samples to the standard curve.
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Fas Associating Death Domain Containing Protein (FADD) were tested 20 times on one plate, respectively Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Fas Associating Death Domain Containing Protein (FADD) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10% Inter-Assay: CV<12%
Restrictions:	For Research Use only

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 2/3 | Product datasheet for ABIN5654058 | 07/25/2024 | Copyright antibodies-online. All rights reserved.

Н	land	lina
11	ana	mig

Handling Advice:	The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.
Storage:	4 °C,-20 °C
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.
Expiry Date:	4-8 months