

# Datasheet for ABIN5654199

# **FGF12 CLIA Kit**



#### Overview

Quantity:	96 tests
Target:	FGF12
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	15.62 pg/mL - 1000 pg/mL
Minimum Detection Limit:	15.62 pg/mL
Application:	ELISA

#### **Product Details**

Sample Type:	Plasma, Serum
Analytical Method:	Quantitative
Detection Method:	Chemiluminescent
Specificity:	This assay has high sensitivity and excellent specificity for detection of Fibroblast Growth Factor 12 (FGF12). No significant cross-reactivity or interference between Fibroblast Growth Factor 12 (FGF12) and analogues was observed.
Sensitivity:	0.53 pg/mL

# Target Details

Target:	FGF12
Alternative Name:	Fibroblast Growth Factor 12 (FGF12 Products)

# Target Details

Background:	Gene Name: Fibroblast Growth Factor 12
	Gene Aliases: FGF12B, FHF1, Fibroblast Growth Factor Homologous Factor 1, Myocyte-
	Activating Factor
Gene ID:	2257
UniProt:	P61328
Pathways:	Negative Regulation of Transporter Activity
Application Details	
Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than
	5 % within the expiration date under appropriate storage condition. To minimize extra influence
	on the performance, operation procedures and lab conditions, especially room temperature, air
	humidity, incubator temperature should be strictly controlled. It is also strongly suggested that
	the whole assay is performed by the same operator from the beginning to the end.
Assay Time:	2-3h
Plate:	Pre-coated
Protocol:	The microplate provided in this kit has been pre-coated with an antibody specific to Fibroblast
	Growth Factor 12 (FGF12). Standards or samples are then added to the appropriate microplate
	wells with a biotin-conjugated antibody specific to Fibroblast Growth Factor 12 (FGF12). Next,
	Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and
	incubated. Then the mixture of substrate A and B is added to generate glow light emission
	kinetics. Upon plate development, the intensity of the emitted light is proportional to the
	Fibroblast Growth Factor 12 (FGF12) level in the sample or standard.,
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level
	Fibroblast Growth Factor 12 (FGF12) were tested 20 times on one plate, respectively
	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level
	Fibroblast Growth Factor 12 (FGF12) were tested on 3 different plates, 8 replicates in each
	plate. CV(%) = SD/meanX100
	Intra-Assay: CV<10%
	Inter-Assay: CV<12%
Restrictions:	For Research Use only

# Handling

Handling Advice:	Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.
Storage:	4 °C,-20 °C
Storage Comment:	-20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.
Expiry Date:	4-8 months