

Datasheet for ABIN5654216 **FGF15 ELISA Kit**



[Go to Product page](#)

Overview

Quantity:	96 tests
Target:	FGF15
Reactivity:	Mouse
Method Type:	Sandwich ELISA
Detection Range:	6.25 pg/mL - 100 pg/mL
Minimum Detection Limit:	6.25 pg/mL
Application:	ELISA

Product Details

Sample Type:	Cell Culture Supernatant, Cell Lysate, Plasma, Serum, Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of Fibroblast Growth Factor 15 (FGF15). No significant cross-reactivity or interference between Fibroblast Growth Factor 15 (FGF15) and analogues was observed.
Sensitivity:	2.52 pg/mL

Target Details

Target:	FGF15
Alternative Name:	Fibroblast Growth Factor 15 (FGF15 Products)

Target Details

Background: Gene Name: Fibroblast Growth Factor 15

Application Details

Comment: The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

Assay Time: 3 h

Plate: Pre-coated

Protocol: This assay employs the competitive inhibition enzyme immunoassay technique. A monoclonal antibody specific to Fibroblast Growth Factor 15 (FGF15) has been pre-coated onto a microplate. A competitive inhibition reaction is launched between biotin labeled Fibroblast Growth Factor 15 (FGF15) and unlabeled Fibroblast Growth Factor 15 (FGF15) (Standards or samples) with the pre-coated antibody specific to Fibroblast Growth Factor 15 (FGF15). After incubation the unbound conjugate is washed off. Next, avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. The amount of bound HRP conjugate is reverse proportional to the concentration of Fibroblast Growth Factor 15 (FGF15) in the sample. After addition of the substrate solution, the intensity of color developed is reverse proportional to the concentration of Fibroblast Growth Factor 15 (FGF15) in the sample.

Assay Precision: Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Fibroblast Growth Factor 15 (FGF15) were tested 20 times on one plate, respectively
Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Fibroblast Growth Factor 15 (FGF15) were tested on 3 different plates, 8 replicates in each plate. $CV(\%) = SD/mean \times 100$
Intra-Assay: $CV < 10\%$
Inter-Assay: $CV < 12\%$

Restrictions: For Research Use only

Handling

Handling Advice: The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.

Handling

Storage: 4 °C,-20 °C

Storage Comment: -20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.

Expiry Date: 4-8 months