

## Datasheet for ABIN5654379 **FOLR1 ELISA Kit**



[Go to Product page](#)

### Overview

Quantity:	96 tests
Target:	FOLR1
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	15.62 pg/mL - 1000 pg/mL
Minimum Detection Limit:	15.62 pg/mL
Application:	ELISA

### Product Details

Sample Type:	Plasma, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of Folate Receptor 1, Adult (FOLR1). No significant cross-reactivity or interference between Folate Receptor 1, Adult (FOLR1) and analogues was observed.
Sensitivity:	5.8 pg/mL

### Target Details

Target:	FOLR1
Alternative Name:	Folate Receptor 1, Adult ( <a href="#">FOLR1 Products</a> )

## Target Details

---

Background:	Gene Name: Folate Receptor 1, Adult Gene Aliases: FBP, FOLR, FR-alpha, MOv18, Adult folate-binding protein, KB cells FBP, Ovarian tumor-associated antigen MOv18
Gene ID:	2348
UniProt:	<a href="#">P15328</a>
Pathways:	<a href="#">Dicarboxylic Acid Transport</a>

## Application Details

---

Comment:	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5 % within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.
Assay Time:	3 h
Plate:	Pre-coated
Protocol:	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Folate Receptor 1, Adult (FOLR1). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Folate Receptor 1, Adult (FOLR1). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Folate Receptor 1, Adult (FOLR1), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Folate Receptor 1, Adult (FOLR1) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Folate Receptor 1, Adult (FOLR1) were tested 20 times on one plate, respectively Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Folate Receptor 1, Adult (FOLR1) were tested on 3 different plates, 8 replicates in each plate. $CV(\%) = SD/mean \times 100$ Intra-Assay: CV<10% Inter-Assay: CV<12%

## Application Details

---

Restrictions: For Research Use only

## Handling

---

Handling Advice: The Stop Solution is acidic. Do not allow to contact skin or eyes. Calibrators, controls and specimen samples should be assayed in duplicate. Once the procedure has been started, all steps should be completed without interruption.

Storage: 4 °C,-20 °C

Storage Comment: -20°C. Bring all reagents to room temperature before beginning test. The kit may be stored at 4°C for immediate use within two days upon arrival. Reseal any unused strips with desiccant pack. Minimize freeze/thaw cycles.

Expiry Date: 4-8 months